

LB 114499


7471B

INSTRUMENT ID: EVL

Linear

 μ Abs.:

48475



A= 0.000

B= 2.048

C= -3.036

Rho= 0.999

Accept=Accept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.000	-0.000	13	0.000	13				
0.2	0.200	0.202	0.002	999	0.0 %	999				
2.5	2.500	2.493	-0.007	12181	0.0 %	12181				
5.0	5.000	4.949	-0.051	24172	0.0 %	24172				
7.5	7.500	7.628	0.128	37248	0.0 %	37248				
10.0	10.000	9.929	-0.071	48475	0.0 %	48475				

Method: 7471B

Operator: Admin

Date of Analysis: 17 May 2021 13:40:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Typ
0.0 - 1	50	13	-	0.00007471B	PPB		17 May 2021 14:22:00	S
0.2 - 1	50.2	999	-	0.20007471B	PPB		17 May 2021 14:27:37	S
2.5 - 1	52.5	12181	-	2.50007471B	PPB		17 May 2021 14:29:54	S
5.0 - 1	55	24172	-	5.00007471B	PPB		17 May 2021 14:32:11	S
7.5 - 1	57.5	37248	-	7.50007471B	PPB		17 May 2021 14:34:27	S
10.0 - 1	510	48475	-	10.00007471B	PPB		17 May 2021 14:40:03	S
ICV30 - 1	ICV30	19641	4.0210	-7471B	PPB		17 May 2021 14:42:49	U
ICB30 - 1	ICB30	-128	-0.0293	-7471B	PPB		17 May 2021 14:45:03	U
CCV12 - 1	CCV12	24558	5.0284	-7471B	PPB		17 May 2021 14:47:19	U
CCB12 - 1	CCB12	-134	-0.0305	-7471B	PPB		17 May 2021 14:49:34	U
CRA - 1	CRA	1023	0.2066	-7471B	PPB		17 May 2021 14:51:50	U
HighStd - 1	HighStd	48299	9.8924	-7471B	PPB		17 May 2021 14:54:05	U
ChkStd - 1	ChkStd	34600	7.0858	-7471B	PPB		17 May 2021 14:56:20	U
PB136480BL - 1	PBS	-171	-0.0381	-7471B	PPB		17 May 2021 14:58:38	U
PB136480BS - 1	LCSS	17794	3.6426	-7471B	PPB		17 May 2021 15:00:54	U
M2404-02 - 1	600494644	2069	0.4209	-7471B	PPB		17 May 2021 15:08:07	U
M2404-02DUP - 1	600494644DUP	2031	0.4131	-7471B	PPB		17 May 2021 15:10:23	U
M2404-02MS - 1	600494644MS	19770	4.0474	-7471B	PPB		17 May 2021 15:12:40	U
M2404-02MSD - 1	600494644MSD	19854	4.0646	-7471B	PPB		17 May 2021 15:14:56	U
M2409-01 - 1	BACKFILL-ASHCROFT-REAR-06	289	0.0562	-7471B	PPB		17 May 2021 15:17:11	U
CCV13 - 1	CCV13	24225	4.9602	-7471B	PPB		17 May 2021 15:19:27	U
CCB13 - 1	CCB13	-59	-0.0151	-7471B	PPB		17 May 2021 15:21:43	U
M2413-01 - 1	124035	7131	1.4580	-7471B	PPB		17 May 2021 15:23:58	U
M2413-03 - 1	124027	10397	2.1271	-7471B	PPB		17 May 2021 15:26:14	U
M2413-05 - 1	124055	2630	0.5358	-7471B	PPB		17 May 2021 15:28:29	U
M2415-01 - 1	124042	896	0.1805	-7471B	PPB		17 May 2021 15:30:45	U
M2404-02LX5 - 1		429	0.0849	-7471B	PPB		17 May 2021 15:33:00	U
M2404-02A - 1		19728	4.0388	-7471B	PPB		17 May 2021 15:35:16	U
CCV14 - 1	CCV14	25966	5.3169	-7471B	PPB		17 May 2021 15:37:32	U
CCB14 - 1	CCB14	-74	-0.0182	-7471B	PPB		17 May 2021 15:39:49	U