



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114502

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	05/17/2021	12:55
2	M2387-03		1	8	100.00	>212.0	>212.0	05/17/2021	13:20
3	M2387-03DUP		1	8	100.00	>212.0	>212.0	05/17/2021	13:45
4	M2387-04		1	9	100.00	>212.0	>212.0	05/17/2021	14:10
5	M2387-06		1	9	100.00	>212.0	>212.0	05/17/2021	14:35
6	M2387-07		1	8	100.00	>212.0	>212.0	05/17/2021	15:05
7	M2387-08		1	9	100.00	>212.0	>212.0	05/17/2021	15:32
8	M2387-09		1	8	100.00	>212.0	>212.0	05/17/2021	15:58
9	M2387-10		1	9	100.00	>212.0	>212.0	05/17/2021	16:25
10	M2388-09		1	9	100.00	>212.0	>212.0	05/17/2021	16:50
11	M2388-11		1	8	100.00	>212.0	>212.0	05/17/2021	17:20
12	M2388-12		1	8	100.00	>212.0	>212.0	05/17/2021	17:48
13	M2405-05		1	9	100.00	>212.0	>212.0	05/17/2021	18:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

kList Name : FP51421-1 WorkList ID : 148939 Department : Wet-Chemistry Date : 05-14-2021 12:21:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
'20/2021	Water	M2387-03	Flash Point	Cool 4 deg C	PSEG03	E22	31055	05/13/2021	1010A
'20/2021	Water	M2387-04	Flash Point	Cool 4 deg C	PSEG03	E22	41102	05/13/2021	1010A
'20/2021	Water	M2387-06	Flash Point	Cool 4 deg C	PSEG03	E22	41799	05/13/2021	1010A
'20/2021	Water	M2387-07	Flash Point	Cool 4 deg C	PSEG03	E22	41800	05/13/2021	1010A
'20/2021	Water	M2387-08	Flash Point	Cool 4 deg C	PSEG03	E22	41801	05/13/2021	1010A
'20/2021	Water	M2387-09	Flash Point	Cool 4 deg C	PSEG03	E22	30794	05/13/2021	1010A
'20/2021	Water	M2387-10	Flash Point	Cool 4 deg C	PSEG03	E22	30795	05/13/2021	1010A
'20/2021	Water	M2388-09	Flash Point	Cool 4 deg C	PSEG03	E22	41113	05/13/2021	1010A
'20/2021	Water	M2388-11	Flash Point	Cool 4 deg C	PSEG03	E22	41101	05/13/2021	1010A
'20/2021	Water	M2388-12	Flash Point	Cool 4 deg C	PSEG03	E22	41769	05/13/2021	1010A
'19/2021	Water	M2405-05	Flash Point	Cool 4 deg C	PSEG03	E31	BUR-1022	05/14/2021	1010A

LB 114502