


LB 114567

7470A

INSTRUMENT ID: CV1

 μ Abs.:

36518



A= 0.0000e+000

B= 2.7983e-004

C= -1.1210e-002

Rho= 0.9991172

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.013	0.013	85	0.000	85					900
0.2	0.200	0.236	0.036	883	0.0 %	883					18
2.5	2.500	2.636	0.136	9459	0.0 %	9459					5
5.0	5.000	4.851	-0.149	17377	0.0 %	17377					-3
7.5	7.500	7.257	-0.243	25973	0.0 %	25973					-3
10.0	10.000	10.208	0.208	36518	0.0 %	36518					2

LB114507 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 May 2021 08:46:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	51	85	-	0.0000	7470A	PPB	18 May 2021 08:52:08	S	Std
0.2 - 1	50.2	883	-	0.2000	7470A	PPB	18 May 2021 08:57:00	S	Std
2.5 - 1	52.5	9459	-	2.5000	7470A	PPB	18 May 2021 08:59:17	S	Std
5.0 - 1	55	17377	-	5.0000	7470A	PPB	18 May 2021 09:01:33	S	Std
7.5 - 1	57.5	25973	-	7.5000	7470A	PPB	18 May 2021 09:06:28	S	Std
10.0 - 1	510	36518	-	10.0000	7470A	PPB	18 May 2021 09:13:53	S	Std
ICV31 - 1	ICV31	13798	3.8499	-	7470A	PPB	18 May 2021 09:16:41	U	SMPL
ICB31 - 1	ICB31	7	-0.0093	-	7470A	PPB	18 May 2021 09:18:56	U	SMPL
CCV15 - 1	CCV15	16690	4.6592	-	7470A	PPB	18 May 2021 09:21:13	U	SMPL
CCB15 - 1	CCB15	-8	-0.0134	-	7470A	PPB	18 May 2021 09:23:29	U	SMPL
CRA - 1	CRA	861	0.2297	-	7470A	PPB	18 May 2021 09:25:44	U	SMPL
HighStd - 1	HighStd	36248	10.1321	-	7470A	PPB	18 May 2021 09:27:59	U	SMPL
ChkStd - 1	ChkStd	24434	6.8262	-	7470A	PPB	18 May 2021 09:30:14	U	SMPL
PB136466BL - 1	PBW	144	0.0291	-	7470A	PPB	18 May 2021 09:32:31	U	SMPL
PB136466BS - 1	LCSW	13343	3.7226	-	7470A	PPB	18 May 2021 09:38:57	U	SMPL
M2405-03 - 1	BUR-1013-1019	250	0.0587	-	7470A	PPB	18 May 2021 09:41:12	U	SMPL
M2407-01 - 1	TEST-PIT-I	302	0.0733	-	7470A	PPB	18 May 2021 09:43:29	U	SMPL
M2407-02 - 1	TEST-PIT-II	320	0.0783	-	7470A	PPB	18 May 2021 09:45:46	U	SMPL
M2413-02 - 1	124035	527	0.1363	-	7470A	PPB	18 May 2021 09:48:03	U	SMPL
M2413-04 - 1	124027	491	0.1262	-	7470A	PPB	18 May 2021 09:50:18	U	SMPL
CCV16 - 1	CCV16	16713	4.6656	-	7470A	PPB	18 May 2021 09:52:33	U	SMPL
CCB16 - 1	CCB16	-13	-0.0148	-	7470A	PPB	18 May 2021 09:54:48	U	SMPL
M2413-06 - 1	124055	369	0.0920	-	7470A	PPB	18 May 2021 09:57:04	U	SMPL
M2415-02 - 1	124042	351	0.0870	-	7470A	PPB	18 May 2021 09:59:19	U	SMPL
M2415-02DUP - 1	124042DUP	333	0.0820	-	7470A	PPB	18 May 2021 10:05:32	U	SMPL
M2415-02MS - 1	124042MS	11862	3.3081	-	7470A	PPB	18 May 2021 10:07:47	U	SMPL
M2415-02MSD - 1	124042MSD	12833	3.5798	-	7470A	PPB	18 May 2021 10:10:03	U	SMPL
PB136467BL - 1	PBW	26	-0.0039	-	7470A	PPB	18 May 2021 10:17:29	U	SMPL
PB136467BS - 1	LCSW	12522	3.4928	-	7470A	PPB	18 May 2021 10:19:44	U	SMPL
M2409-02 - 1	BACKFILL-ASHCROFT-REAR-06	587	0.1531	-	7470A	PPB	18 May 2021 10:26:47	U	SMPL
M2409-02DUP - 1	BACKFILL-ASHCROFT-REAR-06DUP	607	0.1586	-	7470A	PPB	18 May 2021 10:32:12	U	SMPL
M2409-02MS - 1	BACKFILL-ASHCROFT-REAR-06MS	13250	3.6965	-	7470A	PPB	18 May 2021 10:34:28	U	SMPL
CCV17 - 1	CCV17	16963	4.7355	-	7470A	PPB	18 May 2021 10:36:43	U	SMPL
CCB17 - 1	CCB17	-30	-0.0196	-	7470A	PPB	18 May 2021 10:38:58	U	SMPL
M2409-02MSD - 1	BACKFILL-ASHCROFT-REAR-06MSD	12493	3.4847	-	7470A	PPB	18 May 2021 10:41:13	U	SMPL
PB136466TB - 1	PB136419TB	260	0.0615	-	7470A	PPB	18 May 2021 10:43:29	U	SMPL
PB136467TB - 1	PB136430TB	329	0.0809	-	7470A	PPB	18 May 2021 10:45:44	U	SMPL
M2415-02LX5 - 1		52	0.0033	-	7470A	PPB	18 May 2021 10:48:00	U	SMPL
M2415-02A - 1		13459	3.7550	-	7470A	PPB	18 May 2021 10:50:16	U	SMPL
M2409-02LX5 - 1		68	0.0078	-	7470A	PPB	18 May 2021 10:52:32	U	SMPL
M2409-02A - 1		13350	3.7245	-	7470A	PPB	18 May 2021 10:54:48	U	SMPL
CCV18 - 1	CCV18	16768	4.6810	-	7470A	PPB	18 May 2021 10:57:05	U	SMPL
CCB18 - 1	CCB18	-48	-0.0246	-	7470A	PPB	18 May 2021 10:59:22	U	SMPL