



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB114508

Reagent/Standard	Lot/Log #
Residual chlorine LOD, 0.05PPM	WP91164
Residual chlorine ICV-LCS, 0.4PPM	WP91163
Chlorine Calibration std, 0.1ppm	WP91158
Chlorine Calibration std, 0.2ppm	WP91159
Chlorine Calibration std, 0.8ppm	WP91161
Chlorine Calibration std, 0.0ppm	WP91157
Chlorine Calibration std, 1.6ppm	WP91162
Residual Chlorine Calibration and CCV std, 0	WP91160
POTASSIUM BROMATE 0.1N 1L	W2449

Intercept: 0.0089

Slope: 1.0345

Regression: 0.999952

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	-0.01		05/18/2021	10:50
2	CAL2	0.1	1	0.000	0.110	0.110	0.10	2	05/18/2021	10:54
3	CAL3	0.2	1	0.000	0.220	0.220	0.20	-2	05/18/2021	10:57
4	CAL4	0.4	1	0.000	0.430	0.430	0.41	-1.7	05/18/2021	11:01
5	CAL5	0.8	1	0.000	0.840	0.840	0.80	-0.4	05/18/2021	11:05
6	CAL6	1.6	1	0.000	1.660	1.660	1.60	0.3	05/18/2021	11:09



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Alexander

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB114508

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4100	0.4100	0.3880	05/18/2021	11:12
2	ICB				1	0.0000	0.0100	0.0100	0.0010	05/18/2021	11:16
3	CCV1			0.4	1	0.0000	0.4200	0.4200	0.3970	05/18/2021	11:20
4	CCB1				1	0.0000	0.0100	0.0100	0.0010	05/18/2021	11:24
5	LB114508BL	50	50		1	0.0000	0.0100	0.0100	0.0010	05/18/2021	11:27
6	LB114508BS	50	50	0.4	1	0.0000	0.4200	0.4200	0.3970	05/18/2021	11:31
7	M2262-09	50	50		1	0.0000	0.0600	0.0600	0.0490	05/18/2021	11:34
8	CCV2			0.4	1	0.0000	0.4300	0.4300	0.4070	05/18/2021	11:38
9	CCB2				1	0.0000	0.0100	0.0100	0.0010	05/18/2021	11:42