

SFAM01.1


LB114514

INSTRUMENT ID: CV2

Linear

 μ Abs.:

40874



A= 0.0000e+000

B= 2.4415e-004 *SLOPE*C= -3.7397e-002 *y-intercept*

Rho= 0.9999471

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	<i>0/01</i>
0.00	0.000	-0.013	-0.013	101	0.000	101					<i>=</i>
0.05	0.050					204					<i>-14</i>
0.20	0.200	0.173	-0.027	860	0.0 %	860					<i>0</i>
2.50	2.500	2.506	0.006	10416	0.0 %	10416					<i>1</i>
5.00	5.000	5.052	0.052	20847	0.0 %	20847					<i>1</i>
7.50	7.500	7.540	0.040	31035	0.0 %	31035					<i>1</i>
10.0	10.000	9.942	-0.058	40874	0.0 %	40874					<i>-1</i>

LB114514
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	101	-		0 SFAM01.1	PPB	5/18/2021 11:18	Std
	0.2 S01	860	-		0.2 SFAM01.1	PPB	5/18/2021 11:20	Std
	2.5 S02	10416	-		2.5 SFAM01.1	PPB	5/18/2021 11:22	Std
	5 S03	20847	-		5 SFAM01.1	PPB	5/18/2021 11:24	Std
	7.5 S04	31035	-		7.5 SFAM01.1	PPB	5/18/2021 11:27	Std
	10 S05	40874	-		10 SFAM01.1	PPB	5/18/2021 11:29	Std
ICV027	ICV027	16843	4.0749 -		SFAM01.1	PPB	5/18/2021 11:32	SMPL
ICB027	ICB027	6	-0.0359 -		SFAM01.1	PPB	5/18/2021 11:34	SMPL
CCV019	CCV019	20712	5.0195 -		SFAM01.1	PPB	5/18/2021 11:36	SMPL
CCB019	CCB019	-10	-0.0398 -		SFAM01.1	PPB	5/18/2021 11:38	SMPL
PB136487BL	PBS487	3	-0.0367 -		SFAM01.1	PPB	5/18/2021 11:41	SMPL
M2247-01	MHOAB5	884	0.1784 -		SFAM01.1	PPB	5/18/2021 11:43	SMPL
M2247-02	MHOAB6	51924	12.64 -		SFAM01.1	PPB	5/18/2021 11:45	SMPL
M2247-03	MHOAB7	68	-0.0208 -		SFAM01.1	PPB	5/18/2021 11:47	SMPL
M2247-04	MHOAB8	2076	0.4695 -		SFAM01.1	PPB	5/18/2021 11:50	SMPL
M2247-05	MHOAB9	2355	0.5376 -		SFAM01.1	PPB	5/18/2021 11:52	SMPL
M2247-06	MHOAC0	1520	0.3337 -		SFAM01.1	PPB	5/18/2021 11:54	SMPL
M2247-07	MHOAC1	965	0.1982 -		SFAM01.1	PPB	5/18/2021 11:57	SMPL
M2247-08	MHOAC2	25266	6.1314 -		SFAM01.1	PPB	5/18/2021 11:59	SMPL
M2247-09	MHOAC4	2207	0.5014 -		SFAM01.1	PPB	5/18/2021 12:01	SMPL
M2247-10	MHOAC4D	2435	0.5571 -		SFAM01.1	PPB	5/18/2021 12:09	SMPL
M2247-11	MHOAC4S	12649	3.0509 -		SFAM01.1	PPB	5/18/2021 12:11	SMPL
M2247-12	MHOAC5	657	0.123 -		SFAM01.1	PPB	5/18/2021 12:13	SMPL
M2247-13	MHOAC6	2299	0.5239 -		SFAM01.1	PPB	5/18/2021 12:16	SMPL
M2247-14	MHOAC7	2487	0.5698 -		SFAM01.1	PPB	5/18/2021 12:18	SMPL
M2247-15	MHOAC8	2667	0.6138 -		SFAM01.1	PPB	5/18/2021 12:20	SMPL
M2247-16	MHOAC9	873	0.1757 -		SFAM01.1	PPB	5/18/2021 12:22	SMPL
M2247-17	MHOAD0	1966	0.4426 -		SFAM01.1	PPB	5/18/2021 12:25	SMPL
M2247-02DLX5	MHOAB6	11109	2.6749 -		SFAM01.1	PPB	5/18/2021 12:27	SMPL
CCV020	CCV020	20308	4.9209 -		SFAM01.1	PPB	5/18/2021 12:29	SMPL
CCB020	CCB020	21	-0.0323 -		SFAM01.1	PPB	5/18/2021 12:31	SMPL