

LB114516


7471B

INSTRUMENT ID: CV 1

Linear

 μ Abs.:

44964



A= 0.0000e+000

B= 2.2212e-004

C= -4.9306e-002

Rho= 0.9999412

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.014	-0.014	158	0.000	158					
0.2	0.200	0.169	-0.031	981	0.0 %	981					
2.5	2.500	2.513	0.013	11537	0.0 %	11537					
5.0	5.000	5.050	0.050	22956	0.0 %	22956					
7.5	7.500	7.544	0.044	34186	0.0 %	34186					
10.0	10.000	9.938	-0.062	44964	0.0 %	44964					

LB114516 INSTRUMENT ID : CV1

Method: 7471B	Operator: Admin	Date of Analysis: 18 May 2021 12:43:26							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	158	-	0.0000	7471B	PPB	18 May 2021 12:44:50	S	Std
0.2 - 1	SD 2	981	-	0.2000	7471B	PPB	18 May 2021 12:47:07	S	Std
2.5 - 1	SD 5	11537	-	2.5000	7471B	PPB	18 May 2021 12:49:24	S	Std
5.0 - 1	SD 5	22956	-	5.0000	7471B	PPB	18 May 2021 12:51:41	S	Std
7.5 - 1	SD 5	34186	-	7.5000	7471B	PPB	18 May 2021 12:53:57	S	Std
10.0 - 1	SD	44964	-	10.0000	7471B	PPB	18 May 2021 12:56:14	S	Std
ICV32 - 1	ICV32	17876	3.9214	-	7471B	PPB	18 May 2021 12:59:05	U	SMPL
ICB32 - 1	ICB32	186	-0.0080	-	7471B	PPB	18 May 2021 13:01:20	U	SMPL
CCV21 - 1	CCV21	23009	5.0615	-	7471B	PPB	18 May 2021 13:03:36	U	SMPL
CCB21 - 1	CCB21	190	-0.0071	-	7471B	PPB	18 May 2021 13:05:50	U	SMPL
CRA - 1	CRA	947	0.1610	-	7471B	PPB	18 May 2021 13:08:06	U	SMPL
HighStd - 1	HighStd	44788	9.8992	-	7471B	PPB	18 May 2021 13:10:21	U	SMPL
ChkStd - 1	ChkStd	32463	7.1615	-	7471B	PPB	18 May 2021 13:12:37	U	SMPL
PB136506BL - 1	PBS	-168	-0.0866	-	7471B	PPB	18 May 2021 13:14:54	U	SMPL
PB136506BS - 1	LCSS	16672	3.6539	-	7471B	PPB	18 May 2021 13:17:11	U	SMPL
M2366-01 - 1	JG-SPOILS-5-11-21-A	6625	1.4223	-	7471B	PPB	18 May 2021 13:19:27	U	SMPL
M2366-02 - 1	JG-SPOILS-5-11-21-B	4090	0.8592	-	7471B	PPB	18 May 2021 13:21:45	U	SMPL
M2366-03 - 1	JG-SPOILS-5-11-21-C	9391	2.0367	-	7471B	PPB	18 May 2021 13:24:03	U	SMPL
M2420-01 - 1	CONCESSION-CP-1	518	0.0658	-	7471B	PPB	18 May 2021 13:26:19	U	SMPL
M2420-01DUP - 1	CONCESSION-CP-1DUP	559	0.0749	-	7471B	PPB	18 May 2021 13:31:34	U	SMPL
CCV22 - 1	CCV22	23203	5.1046	-	7471B	PPB	18 May 2021 13:33:48	U	SMPL
CCB22 - 1	CCB22	215	-0.0015	-	7471B	PPB	18 May 2021 13:36:03	U	SMPL
M2420-01MS - 1	CONCESSION-CP-1MS	18288	4.0129	-	7471B	PPB	18 May 2021 13:38:19	U	SMPL
M2420-01MSD - 1	CONCESSION-CP-1MSD	17071	3.7426	-	7471B	PPB	18 May 2021 13:40:34	U	SMPL
M2420-02 - 1	CONCESSION-CP-2	106	-0.0258	-	7471B	PPB	18 May 2021 13:42:56	U	SMPL
M2420-03 - 1	CONCESSION-CP-3	630	0.0906	-	7471B	PPB	18 May 2021 13:45:13	U	SMPL
M2420-04 - 1	CONCESSION-CP-4	572	0.0777	-	7471B	PPB	18 May 2021 13:47:29	U	SMPL
M2420-05 - 1	CONCESSION-CP-5	131	-0.0202	-	7471B	PPB	18 May 2021 13:49:45	U	SMPL
M2420-06 - 1	PLAYGROUND-CP-1	3232	0.6686	-	7471B	PPB	18 May 2021 13:52:01	U	SMPL
M2433-01 - 1	GPLD	1878	0.3678	-	7471B	PPB	18 May 2021 13:54:18	U	SMPL
M2435-01 - 1	VAN-REIPEN-A	5561	1.1859	-	7471B	PPB	18 May 2021 13:56:34	U	SMPL
M2435-04 - 1	VAN-REIPEN-B	1138	0.2035	-	7471B	PPB	18 May 2021 13:58:50	U	SMPL
CCV23 - 1	CCV23	22851	5.0264	-	7471B	PPB	18 May 2021 14:01:05	U	SMPL
CCB23 - 1	CCB23	125	-0.0215	-	7471B	PPB	18 May 2021 14:03:20	U	SMPL
M2435-07 - 1	HOMESTEAD-A	3629	0.7568	-	7471B	PPB	18 May 2021 14:05:36	U	SMPL
M2435-10 - 1	HOMESTEAD-B	158	-0.0142	-	7471B	PPB	18 May 2021 14:07:52	U	SMPL
M2435-13 - 1	COTTAGE-A	10324	2.2439	-	7471B	PPB	18 May 2021 14:10:07	U	SMPL
M2435-16 - 1	COTTAGE-B	908	0.1524	-	7471B	PPB	18 May 2021 14:12:23	U	SMPL
M2436-01 - 1	TP-1A	13976	3.0551	-	7471B	PPB	18 May 2021 14:14:38	U	SMPL
M2436-04 - 1	TP-2A	12259	2.6737	-	7471B	PPB	18 May 2021 14:16:54	U	SMPL
M2436-07 - 1	TP-3A	27792	6.1240	-	7471B	PPB	18 May 2021 14:19:11	U	SMPL
M2420-01LX5 - 1		114	-0.0240	-	7471B	PPB	18 May 2021 14:21:28	U	SMPL
M2420-01A - 1		17692	3.8805	-	7471B	PPB	18 May 2021 14:23:45	U	SMPL
CCV24 - 1	CCV24	22638	4.9791	-	7471B	PPB	18 May 2021 14:26:00	U	SMPL
CCB24 - 1	CCB24	202	-0.0044	-	7471B	PPB	18 May 2021 14:28:15	U	SMPL