

LB114517

Test results

Aquakem 7.2AQ1

Page: 1

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092Reviewed by : AM Instrument ID : Konelab

5/18/2021 13:08

Test: Phenolics-

Sample Id	Result	Dil. 1 +	Response	Errors
ICV1	0.993	0.0	0.104	
ICB1	0.005	0.0	0.005	
CCV1	0.990	0.0	0.103	
CCB1	0.006	0.0	0.006	
PB136508BL	0.006	0.0	0.006	
PB136508BS	0.985	0.0	0.103	
M2262-12	0.029	0.0	0.008	
CCV2	0.989	0.0	0.103	
CCB2	0.006	0.0	0.006	

N	9
Mean	0.445
SD	0.5157
CV%	115.78

Aquakem v. 7.2AQ1

Results from time period:

Tue May 18 12:12:09 2021

Tue May 18 13:07:33 2021

Sample Id	Sam/Ctr/c/	Test short	Test type	Result	Result unit	Result date and time	Stat
0.0PPM	A	Phenolics-	P	0.0027	mg/l	5/18/2021 12:12:09	
0.05PPM	A	Phenolics-	P	0.0521	mg/l	5/18/2021 12:12:10	
0.1PPM	A	Phenolics-	P	0.1009	mg/l	5/18/2021 12:12:11	
0.25PPM	A	Phenolics-	P	0.2511	mg/l	5/18/2021 12:12:12	
0.50PPM	A	Phenolics-	P	0.5045	mg/l	5/18/2021 12:12:13	
1.0PPM	A	Phenolics-	P	0.98	mg/l	5/18/2021 12:12:14	
2.0PPM	A	Phenolics-	P	2.0086	mg/l	5/18/2021 12:12:15	
ICV1	S	Phenolics-	P	0.9927	mg/l	5/18/2021 13:01:39	
ICB1	S	Phenolics-	P	0.0054	mg/l	5/18/2021 13:01:40	
CCV1	S	Phenolics-	P	0.9896	mg/l	5/18/2021 13:01:43	
CCB1	S	Phenolics-	P	0.0062	mg/l	5/18/2021 13:01:45	
PB136508BL	S	Phenolics-	P	0.0064	mg/l	5/18/2021 13:01:46	
PB136508BS	S	Phenolics-	P	0.9846	mg/l	5/18/2021 13:01:48	
M2262-12	S	Phenolics-	P	0.0289	mg/l	5/18/2021 13:07:29	
CCV2	S	Phenolics-	P	0.989	mg/l	5/18/2021 13:07:31	
CCB2	S	Phenolics-	P	0.0061	mg/l	5/18/2021 13:07:33	

CHEMTECH CONSULTING GROUP INC
284 Sheffield Street, Mountainside, NJ 07092

Reviewed by : AM Instrument ID : Konelab

5/18/2021 12:12


Test Phenolics-

Accepted 5/18/2021 12:12

Factor 10.06
Bias 0.005

Coeff. of det. 0.999839

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors %RE
1	0.00PPM	0.005	0.0027	0.0000	4.2
2	Phenol-2	0.010	0.0521	0.0500	0.9
3	Phenol-2	0.015	0.1009	0.1000	0.4
4	Phenol-2	0.030	0.2511	0.2500	0.9
5	Phenol-2	0.055	0.5045	0.5000	0.9
6	Phenol-2	0.102	0.9800	1.0000	0.4
7	Phenol-2	0.204	2.0086	2.0000	

05/18/2021
AM