



TOTAL VOLATILE SOLIDS SM2540 G

TEMP1 IN: 104 °C 05/18/2021 16:20 TEMP1 OUT: 103 °C 05/19/2021 08:00
TEMP2 IN: 103 °C 05/19/2021 08:30 TEMP2 OUT: 103 °C 05/19/2021 10:00
TEMP3 IN: 540 °C 05/19/2021 10:30 TEMP3 OUT: 540 °C 05/19/2021 12:00
TEMP4 IN: 560 °C 05/19/2021 12:30 TEMP4 OUT: 560 °C 05/19/2021 14:00

Run Number: LB114523

SUPERVISOR: apatel

ANALYST: jignesh

BalanceID: WC SC-3

OvenID: WC SC-3

Dish #	Lab ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish + SampleWt Drying @103-@105°C (g)	Final Dish + SampleWt Drying @103-@105°C (g)	Dish + Samplewt Drying @550 (±50) °C (g)	Final Dish + Samplewt Drying @550 (±50) °C (g)	Weight Diff (g)	Result (mg/L)
1	LB114523BL	79.8637	79.8637	100	79.8638	79.8638	79.8638	79.8638	0.0000	0
2	M2440-01	89.9931	89.9931	100	90.0875	90.0875	90.0710	90.0710	0.0165	165
3	M2440-01DUP	90.7201	90.7201	100	90.8142	90.8142	90.7985	90.7985	0.0157	157

A = Sample Volume (ml)

B = Final Dish + Samplewt Drying @550 (±50) °C (g)

C = Final Dish + SampleWt Drying @103-@105°C (g)

D = Weight (g)

E = Final Empty Dish Weight (g)

F = Final Dish + SampleWt Drying @103-@105°C (g)

Weight D = C - B

Result % = $\frac{D}{A} \times 1000000$

JP 114523

WORKLIST(Hardcopy Internal Chain)

WorkList Name: tvs m2440 w WorkList ID: 149037 Department: Wet-Chemistry Date: 05-18-2021 15:52:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/25/2021	Water	M2440-01	L TVS	Cool 4 deg C	PSEG03	E41	TOTE-1	05/18/2021	SM2540 G

Date/Time 05-18-21 16:00
Raw Sample Received by: JP (Wet Chem)
Raw Sample Relinquished by: JP SM

Date/Time 05-18-21 17:00
Raw Sample Received by: JP SM
Raw Sample Relinquished by: JP (Wet Chem)