



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114524

Slope : 99.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/18/2021	16:10
2	CAL2	1	Water	NA	NA	21.2	7.00	05/18/2021	16:11
3	CAL3	1	Water	NA	NA	21.3	10.04	05/18/2021	16:12
4	ICV	1	Water	NA	NA	21.3	7.02	05/18/2021	16:13
5	CCV1	1	Water	NA	NA	21.4	2.01	05/18/2021	16:15
6	M2440-01	1	Water	NA	NA	25.6	6.97	05/18/2021	16:16
7	M2440-01DUP	1	Water	NA	NA	25.7	6.98	05/18/2021	16:17
8	CCV2	1	Water	NA	NA	21.5	12.02	05/18/2021	16:19

2B114524

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2440 WorkList ID : 149038 Department : Wet-Chemistry Date : 05-18-2021 15:53:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/18/2021	Water	M2440-01	pH	Cool 4 deg C	PSEG03	E41	TOTE-1	05/18/2021	9040C

Date/Time 05/18/21 16:00
Raw Sample Received by: JP Wet Chem
Raw Sample Relinquished by: CP SM

Date/Time 05/18/21 17:00
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JP Wet Chem