



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114525

Slope : 99.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/18/2021	16:10
2	CAL2	1	Water	NA	NA	21.2	7.00	05/18/2021	16:11
3	CAL3	1	Water	NA	NA	21.3	10.04	05/18/2021	16:12
4	ICV	1	Water	NA	NA	21.3	7.02	05/18/2021	16:13
5	CCV1	1	Water	NA	NA	21.4	2.01	05/18/2021	16:15
6	M2437-02	1	Solid	20.02	20	24.3	8.47	05/18/2021	16:20
7	M2437-02DUP	1	Solid	20.03	20	24.4	8.48	05/18/2021	16:21
8	M2437-05	1	Solid	20.04	20	24.4	8.98	05/18/2021	16:23
9	M2437-08	1	Solid	20.02	20	24.0	8.70	05/18/2021	16:25
10	CCV2	1	Water	NA	NA	20.3	12.02	05/18/2021	16:30

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2437 WorkList ID : 149035 Department : Wet-Chemistry Date : 05-18-2021 15:31:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/18/2021	Solid	M2437-02	Corrosivity	Cool 4 deg C	PSEG03	E41	TP-1	05/18/2021	9045D
05/18/2021	Solid	M2437-05	Corrosivity	Cool 4 deg C	PSEG03	E41	TP-2	05/18/2021	9045D
05/18/2021	Solid	M2437-08	Corrosivity	Cool 4 deg C	PSEG03	E41	TP-3	05/18/2021	9045D

Date/Time 05-18-21 16:00
 Raw Sample Received by: JP
 Raw Sample Relinquished by: JP SM

Date/Time 05-18-21 17:20
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JP Ret

chem

28 114525