



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114537

Slope : 99.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	05/19/2021	11:20
2	CAL2	1	Water	NA	NA	21.4	7.01	05/19/2021	11:21
3	CAL3	1	Water	NA	NA	21.4	10.04	05/19/2021	11:22
4	ICV	1	Water	NA	NA	21.3	6.98	05/19/2021	11:25
5	CCV1	1	Water	NA	NA	21.4	2.01	05/19/2021	11:29
6	M2454-01	1	Water	NA	NA	23.7	6.69	05/19/2021	11:30
7	M2454-01DUP	1	Water	NA	NA	23.6	6.70	05/19/2021	11:31
8	CCV2	1	Water	NA	NA	21.5	12.02	05/19/2021	11:32

2B 114537

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2454 WorkList ID : 149079 Department : Wet-Chemistry Date : 05-19-2021 11:16:32

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/14/2021	Water	M2454-01	pH	Cool 4 deg C	STER06	F21	POOL-WATER	05/14/2021	9040C

Date/Time 05-19-21 11:20

Raw Sample Received by: JP (Wf Chem)

Raw Sample Relinquished by: CJP SM

Date/Time 05-19-21 12:10

Raw Sample Received by: CJP SM

Raw Sample Relinquished by: JP (Wf Chem)