


LB114549

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.6270e-004 slope

C= -1.8026e-002 y-intercept

Rho= 0.9999486

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/v
0.00	0.000	-0.016	-0.016	6	0.000	6					
0.05	0.050					204					
0.20	0.200	0.178	-0.022	746	0.0 %	746					
2.50	2.500	2.500	-0.000	9585	0.0 %	9585					
5.00	5.000	5.078	0.078	19400	0.0 %	19400					
7.50	7.500	7.499	-0.001	28615	0.0 %	28615					
10.0	10.000	9.961	-0.039	37987	0.0 %	37987					

o/v

11

0020

LB114549
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	6	-		0 SFAM01.1	PPB	5/19/2021 15:38	Std
	0.2 S01	746	-		0.2 SFAM01.1	PPB	5/19/2021 15:40	Std
	2.5 S02	9585	-		2.5 SFAM01.1	PPB	5/19/2021 15:43	Std
	5 S03	19400	-		5 SFAM01.1	PPB	5/19/2021 15:45	Std
	7.5 S04	28615	-		7.5 SFAM01.1	PPB	5/19/2021 15:47	Std
	10 S05	37987	-		10 SFAM01.1	PPB	5/19/2021 15:52	Std
ICV028	ICV028	15555	4.0683 -		SFAM01.1	PPB	5/19/2021 15:56	SMPL
ICB028	ICB028	-122	-0.0501 -		SFAM01.1	PPB	5/19/2021 15:58	SMPL
CCV027	CCV027	19385	5.0744 -		SFAM01.1	PPB	5/19/2021 16:00	SMPL
CCB027	CCB027	-123	-0.0503 -		SFAM01.1	PPB	5/19/2021 16:02	SMPL
PB136529BL	PBW529	-14	-0.0217 -		SFAM01.1	PPB	5/19/2021 16:05	SMPL
M2446-01	MBG1Z3	283	0.0563 -		SFAM01.1	PPB	5/19/2021 16:07	SMPL
M2446-02	MBG1Z4	19	-0.013 -		SFAM01.1	PPB	5/19/2021 16:09	SMPL
M2446-03	MBG1Z4D	30	-0.0101 -		SFAM01.1	PPB	5/19/2021 16:11	SMPL
M2446-04	MBG1Z4S	4385	1.1339 -		SFAM01.1	PPB	5/19/2021 16:14	SMPL
PB136531BL	PBW531	-25	-0.0246 -		SFAM01.1	PPB	5/19/2021 16:16	SMPL
M2447-01	MBG229	253	0.0484 -		SFAM01.1	PPB	5/19/2021 16:18	SMPL
M2447-02	MBG230	12	-0.0149 -		SFAM01.1	PPB	5/19/2021 16:21	SMPL
M2447-03	MBG230D	13	-0.0146 -		SFAM01.1	PPB	5/19/2021 16:23	SMPL
M2447-04	MBG230S	3946	1.0186 -		SFAM01.1	PPB	5/19/2021 16:25	SMPL
CCV028	CCV028	19389	5.0754 -		SFAM01.1	PPB	5/19/2021 16:27	SMPL
CCB028	CCB028	-104	-0.0453 -		SFAM01.1	PPB	5/19/2021 16:30	SMPL