



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114552

Slope : 99.8

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.5	4.01	05/19/2021	16:00
2	CAL2	1	Water	NA	NA	21.5	7.00	05/19/2021	16:01
3	CAL3	1	Water	NA	NA	21.5	10.02	05/19/2021	16:02
4	ICV	1	Water	NA	NA	21.5	6.98	05/19/2021	16:05
5	CCV1	1	Water	NA	NA	21.6	2.01	05/19/2021	16:07
6	M2457-02	1	Solid	20.02	20	24.4	8.36	05/19/2021	16:10
7	M2457-02DUP	1	Solid	20.03	20	24.5	8.39	05/19/2021	16:11
8	M2467-01	1	Solid	20.02	20	23.0	6.65	05/19/2021	16:13
9	CCV2	1	Water	NA	NA	21.4	12.02	05/19/2021	16:15

114532

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY M2467 WorkList ID : 149087 Department : Wet-Chemistry Date : 05-19-2021 15:43:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2021	Solid	M2457-02	Corrosivity	Cool 4 deg C	PSEG03	E42	TP-4	05/19/2021	9045D
05/19/2021	Solid	M2467-01	Corrosivity	Cool 4 deg C	JACO05	F11	IDW-BR3-SO-051921	05/19/2021	9045D

Date/Time 05.19.21 1545
 Raw Sample Received by: JO (wet Chem)
 Raw Sample Relinquished by: KS (wet Chem)

Date/Time 05.19.21 1640
 Raw Sample Received by: KS (wet Chem)
 Raw Sample Relinquished by: JO (wet Chem)