

LB11455.5


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

36502



A= 0.0000e+000

B= 2.7388e-004

C= -1.1013e-002

Rho= 0.9999602

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.0	0.000	-0.002	-0.002	32	0.000	32					
0.2	0.200	0.166	-0.034	647	0.0 %	647					-17
2.5	2.500	2.515	0.015	9224	0.0 %	9224					1
5.0	5.000	5.064	0.064	18529	0.0 %	18529					1
7.5	7.500	7.471	-0.029	27317	0.0 %	27317					0
10.0	10.000	9.986	-0.014	36502	0.0 %	36502					0

LB114555 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 20 May 2021 08:56:28					
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	7470A	PPB	20 May 2021 09:07:34	S	Std
0.2 - 1	50.2	647	-	0.2000	7470A	PPB	20 May 2021 09:09:51	S	Std
2.5 - 1	52.5	9224	-	2.5000	7470A	PPB	20 May 2021 09:12:08	S	Std
5.0 - 1	55	18529	-	5.0000	7470A	PPB	20 May 2021 09:14:25	S	Std
7.5 - 1	57.5	27317	-	7.5000	7470A	PPB	20 May 2021 09:16:41	S	Std
10.0 - 1	51V	36502	-	10.0000	7470A	PPB	20 May 2021 09:21:43	S	Std
ICV34 - 1	ICV34	13970	3.8151	-	7470A	PPB	20 May 2021 09:24:35	U	SMPL
ICB34 - 1	ICB34	-23	-0.0173	-	7470A	PPB	20 May 2021 09:29:22	U	SMPL
CCV29 - 1	CCV29	18401	5.0287	-	7470A	PPB	20 May 2021 09:31:37	U	SMPL
CCB29 - 1	CCB29	-6	-0.0127	-	7470A	PPB	20 May 2021 09:36:53	U	SMPL
CRA - 1	CRA	585	0.1492	-	7470A	PPB	20 May 2021 09:41:45	U	SMPL
HighStd - 1	HighStd	35933	9.8305	-	7470A	PPB	20 May 2021 09:44:20	U	SMPL
ChkStd - 1	ChkStd	25205	6.8922	-	7470A	PPB	20 May 2021 09:46:35	U	SMPL
PB136543BL - 1	PBW	-217	-0.0704	-	7470A	PPB	20 May 2021 09:48:53	U	SMPL
PB136543BS - 1	LCSW	13852	3.7828	-	7470A	PPB	20 May 2021 09:51:09	U	SMPL
M2387-12 - 1	41789	-117	-0.0431	-	7470A	PPB	20 May 2021 09:53:25	U	SMPL
M2437-03 - 1	TP-1	2	-0.0105	-	7470A	PPB	20 May 2021 09:55:42	U	SMPL
M2437-06 - 1	TP-2	20	-0.0055	-	7470A	PPB	20 May 2021 09:57:59	U	SMPL
M2437-09 - 1	TP-3	25	-0.0042	-	7470A	PPB	20 May 2021 10:00:16	U	SMPL
M2437-09DUP - 1	TP-3DUP	27	-0.0036	-	7470A	PPB	20 May 2021 10:02:31	U	SMPL
CCV30 - 1	CCV30	18093	4.9444	-	7470A	PPB	20 May 2021 10:04:47	U	SMPL
CCB30 - 1	CCB30	-152	-0.0526	-	7470A	PPB	20 May 2021 10:07:02	U	SMPL
M2437-09MS - 1	TP-3MS	13079	3.5711	-	7470A	PPB	20 May 2021 10:09:18	U	SMPL
M2437-09MSD - 1	TP-3MSD	13473	3.6790	-	7470A	PPB	20 May 2021 10:11:33	U	SMPL
PB136544BL - 1	PBW	-115	-0.0425	-	7470A	PPB	20 May 2021 10:13:48	U	SMPL
PB136544BS - 1	LCSW	15512	4.2375	-	7470A	PPB	20 May 2021 10:16:04	U	SMPL
M2354-05 - 1	EO-01-051121	-134	-0.0477	-	7470A	PPB	20 May 2021 10:18:20	U	SMPL
M2366-07 - 1	JG-SPOILS-5-11-21-A	7	-0.0091	-	7470A	PPB	20 May 2021 10:20:36	U	SMPL
M2366-08 - 1	JG-SPOILS-5-11-21-B	-3	-0.0118	-	7470A	PPB	20 May 2021 10:22:52	U	SMPL
M2366-09 - 1	JG-SPOILS-5-11-21-C	-6	-0.0127	-	7470A	PPB	20 May 2021 10:25:09	U	SMPL
M2435-02 - 1	VAN-REIPEN-A	6	-0.0094	-	7470A	PPB	20 May 2021 10:27:25	U	SMPL
M2435-05 - 1	VAN-REIPEN-B	30	-0.0028	-	7470A	PPB	20 May 2021 10:29:40	U	SMPL
CCV31 - 1	CCV31	17789	4.8611	-	7470A	PPB	20 May 2021 10:31:56	U	SMPL
CCB31 - 1	CCB31	-138	-0.0488	-	7470A	PPB	20 May 2021 10:34:11	U	SMPL
M2435-08 - 1	HOMESTEAD-A	6	-0.0094	-	7470A	PPB	20 May 2021 10:36:26	U	SMPL
M2435-11 - 1	HOMESTEAD-B	12	-0.0077	-	7470A	PPB	20 May 2021 10:38:41	U	SMPL
M2435-14 - 1	COTTAGE-A	124	0.0229	-	7470A	PPB	20 May 2021 10:40:57	U	SMPL
M2435-17 - 1	COTTAGE-B	26	-0.0039	-	7470A	PPB	20 May 2021 10:43:12	U	SMPL
M2436-02 - 1	TP-1A	21	-0.0053	-	7470A	PPB	20 May 2021 10:45:28	U	SMPL
M2436-05 - 1	TP-2A	19	-0.0058	-	7470A	PPB	20 May 2021 10:47:44	U	SMPL
M2436-08 - 1	TP-3A	41	0.0002	-	7470A	PPB	20 May 2021 10:50:00	U	SMPL
M2436-08DUP - 1	TP-3ADUP	21	-0.0053	-	7470A	PPB	20 May 2021 10:52:17	U	SMPL
M2436-08MS - 1	TP-3AMS	12858	3.5106	-	7470A	PPB	20 May 2021 10:54:34	U	SMPL
M2436-08MSD - 1	TP-3AMSD	13918	3.8009	-	7470A	PPB	20 May 2021 10:56:50	U	SMPL
CCV32 - 1	CCV32	17726	4.8439	-	7470A	PPB	20 May 2021 11:01:38	U	SMPL
CCB32 - 1	CCB32	-145	-0.0507	-	7470A	PPB	20 May 2021 11:03:54	U	SMPL
PB136543TB - 1	PB136504TB	-13	-0.0146	-	7470A	PPB	20 May 2021 11:06:10	U	SMPL
PB136544TB - 1	PB136464TB	13	-0.0075	-	7470A	PPB	20 May 2021 11:08:26	U	SMPL
M2437-09LX5 - 1		1	-0.0107	-	7470A	PPB	20 May 2021 11:10:43	U	SMPL
M2437-09A - 1		13661	3.7305	-	7470A	PPB	20 May 2021 11:12:58	U	SMPL
M2436-08LX5 - 1		-105	-0.0398	-	7470A	PPB	20 May 2021 11:15:15	U	SMPL
M2436-08A - 1		13682	3.7363	-	7470A	PPB	20 May 2021 11:17:31	U	SMPL
CCV33 - 1	CCV33	17739	4.8474	-	7470A	PPB	20 May 2021 11:19:48	U	SMPL
CCB33 - 1	CCB33	-142	-0.0499	-	7470A	PPB	20 May 2021 11:22:05	U	SMPL