



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114563

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	05/19/2021	09:45
2	M2440-01		1	9	100.00	>212.0	>212.0	05/19/2021	10:10
3	M2440-01DUP		1	9	100.00	>212.0	>212.0	05/19/2021	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 114563

WorkList Name : fp51921 WorkList ID : 149065 Department : Wet-Chemistry Date : 05-19-2021 09:06:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/25/2021	Water	M2440-01	Flash Point	Cool 4 deg C	PSEG03	E41	TOTE-1	05/18/2021	1010A

Date/Time 05/19/21 09:15
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: CL SM

Date/Time 05/19/21 10:45
 Raw Sample Received by: CL SM
 Raw Sample Relinquished by: KS (Wetchem)