



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114570

Slope : 99.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	05/20/2021	10:38
2	CAL2	1	Water	NA	NA	21.3	7.00	05/20/2021	10:39
3	CAL3	1	Water	NA	NA	21.3	10.02	05/20/2021	10:40
4	ICV	1	Water	NA	NA	21.3	7.01	05/20/2021	10:41
5	CCV1	1	Water	NA	NA	21.4	2.01	05/20/2021	10:42
6	M2471-01	1	Water	NA	NA	22.8	5.74	05/20/2021	10:45
7	M2471-01DUP	1	Water	NA	NA	22.9	5.75	05/20/2021	10:46
8	M2471-02	1	Water	NA	NA	23.1	5.35	05/20/2021	10:47
9	M2471-03	1	Water	NA	NA	23.2	5.26	05/20/2021	10:49
10	CCV2	1	Water	NA	NA	21.5	12.02	05/20/2021	10:50

2B 114510

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2471 WorkList ID : 149102 Department : Wet-Chemistry Date : 05-20-2021 10:32:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/19/2021	Water	M2471-01	pH	Cool 4 deg C	STER13	F11	CELL-A	05/19/2021	9040C
05/19/2021	Water	M2471-02	pH	Cool 4 deg C	STER13	F11	CELL-B	05/19/2021	9040C
05/19/2021	Water	M2471-03	pH	Cool 4 deg C	STER13	F11	CELL-C	05/19/2021	9040C

Date/Time 05/20/21 10:35
Raw Sample Received by: JO (wet Chem)
Raw Sample Relinquished by: CP SM

Date/Time 05/20/21 11:40
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JO (wet Chem)