


LB114574

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5871e-004 *slope*C= -2.8149e-005 *intercept*

Rho= 0.9999762

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.010	0.010	38	0.000	38					1
0.05	0.050					204					2
0.20	0.200	0.197	-0.003	760	0.0 %	760					0
2.50	2.500	2.500	0.000	9665	0.0 %	9665					0
5.00	5.000	5.013	0.013	19376	0.0 %	19376					0
7.50	7.500	7.449	-0.051	28792	0.0 %	28792					1
10.0	10.000	10.032	0.032	38775	0.0 %	38775					0

LB114574
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 SFAM01.1	PPB	5/20/2021 15:03	Std
	0.2 S01	760	-		0.2 SFAM01.1	PPB	5/20/2021 15:05	Std
	2.5 S02	9665	-		2.5 SFAM01.1	PPB	5/20/2021 15:08	Std
	5 S03	19376	-		5 SFAM01.1	PPB	5/20/2021 15:10	Std
	7.5 S04	28792	-		7.5 SFAM01.1	PPB	5/20/2021 15:12	Std
	10 S05	38775	-		10 SFAM01.1	PPB	5/20/2021 15:14	Std
ICV029	ICV029	15319	3.9632 -		SFAM01.1	PPB	5/20/2021 15:19	SMPL
ICB029	ICB029	-96	-0.0249 -		SFAM01.1	PPB	5/20/2021 15:21	SMPL
CCV038	CCV038	18740	4.8482 -		SFAM01.1	PPB	5/20/2021 15:23	SMPL
CCB038	CCB038	-103	-0.0267 -		SFAM01.1	PPB	5/20/2021 15:25	SMPL
PB136571BL	PBW571	11	0.0028 -		SFAM01.1	PPB	5/20/2021 15:28	SMPL
M2286-01	MHOAH4	234	0.0605 -		SFAM01.1	PPB	5/20/2021 15:30	SMPL
M2286-02	MHOAH6	47	0.0121 -		SFAM01.1	PPB	5/20/2021 15:32	SMPL
M2286-03	MHOAH8	64	0.0165 -		SFAM01.1	PPB	5/20/2021 15:34	SMPL
M2286-04	MHOAJ0	48	0.0124 -		SFAM01.1	PPB	5/20/2021 15:37	SMPL
M2286-05	MHOAJ2	65	0.0168 -		SFAM01.1	PPB	5/20/2021 15:39	SMPL
M2286-06	MHOAJ4	63	0.0163 -		SFAM01.1	PPB	5/20/2021 15:41	SMPL
M2286-07	MHOAJ6	114	0.0295 -		SFAM01.1	PPB	5/20/2021 15:43	SMPL
M2286-08	MHOAJ8	226	0.0584 -		SFAM01.1	PPB	5/20/2021 15:46	SMPL
M2286-09	MHOAP1	60	0.0155 -		SFAM01.1	PPB	5/20/2021 15:48	SMPL
M2286-10	MHOAP3	751	0.1943 -		SFAM01.1	PPB	5/20/2021 15:50	SMPL
M2286-11	MHOAP5	1885	0.4876 -		SFAM01.1	PPB	5/20/2021 15:53	SMPL
CCV039	CCV039	19014	4.9191 -		SFAM01.1	PPB	5/20/2021 15:55	SMPL
CCB039	CCB039	-108	-0.028 -		SFAM01.1	PPB	5/20/2021 15:57	SMPL
M2286-12	MHOAP7	604	0.1562 -		SFAM01.1	PPB	5/20/2021 16:00	SMPL
M2286-13	MHOAP7D	599	0.1549 -		SFAM01.1	PPB	5/20/2021 16:02	SMPL
M2286-14	MHOAP7S	4172	1.0793 -		SFAM01.1	PPB	5/20/2021 16:04	SMPL
M2286-15	MHOAP9	24	0.0062 -		SFAM01.1	PPB	5/20/2021 16:07	SMPL
M2286-16	MHOAQ1	75	0.0194 -		SFAM01.1	PPB	5/20/2021 16:09	SMPL
M2286-17	MHOAU3	16	0.0041 -		SFAM01.1	PPB	5/20/2021 16:11	SMPL
M2286-18	MHOAU4	19	0.0049 -		SFAM01.1	PPB	5/20/2021 16:14	SMPL
M2286-19	MHOAU5	29	0.0075 -		SFAM01.1	PPB	5/20/2021 16:16	SMPL
M2286-20	MHOAU6	54	0.0139 -		SFAM01.1	PPB	5/20/2021 16:18	SMPL
M2286-21	MHOAU7	164	0.0424 -		SFAM01.1	PPB	5/20/2021 16:20	SMPL
CCV040	CCV040	19365	5.0099 -		SFAM01.1	PPB	5/20/2021 16:23	SMPL
CCB040	CCB040	-95	-0.0246 -		SFAM01.1	PPB	5/20/2021 16:25	SMPL