



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114576

Slope : 99.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.4	4.01	05/20/2021	16:15
2	CAL2	1	Water	NA	NA	21.4	7.01	05/20/2021	16:16
3	CAL3	1	Water	NA	NA	21.4	10.03	05/20/2021	16:17
4	ICV	1	Water	NA	NA	21.4	6.99	05/20/2021	16:18
5	CCV1	1	Water	NA	NA	21.4	2.01	05/20/2021	16:20
6	M2482-01	1	Solid	20.03	20	24.9	11.35	05/20/2021	16:23
7	M2482-01DUP	1	Solid	20.04	20	24.8	11.36	05/20/2021	16:24
8	CCV2	1	Water	NA	NA	21.5	12.03	05/21/2021	16:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2477 WorkList ID : 149114 Department : Wet-Chemistry Date : 05-20-2021 15:57:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/20/2021	Solid	M2482-01	pH	Cool 4 deg C	PSEG05	E41	BU-02-052021	05/20/2021	9045D

Date/Time 05/20/21 15:00
 Raw Sample Received by: JR (Signature)
 Raw Sample Relinquished by: L.C. (Signature)

Date/Time 05/20/21 17:00
 Raw Sample Received by: L.C. (Signature)
 Raw Sample Relinquished by: JR (Signature)