

LB114578


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

37748



A= 0.0000e+000

B=

2.6687e-004

slope

C=

1.2369e-003

y-intercept

Rho=

0.9999018

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.045	0.045	164	0.000	164					11
0.05	0.050					204					5
0.20	0.200	0.209	0.009	779	0.0 %	779					0
2.50	2.500	2.498	-0.002	9354	0.0 %	9354					2
5.00	5.000	4.925	-0.075	18449	0.0 %	18449					1
7.50	7.500	7.449	-0.051	27907	0.0 %	27907					1
10.0	10.000	10.075	0.075	37748	0.0 %	37748					1

LB114578
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	164	-		0 SFAM01.1	PPB	5/21/2021 8:45	Std
	0.2 S01	779	-		0.2 SFAM01.1	PPB	5/21/2021 8:48	Std
	2.5 S02	9354	-		2.5 SFAM01.1	PPB	5/21/2021 8:50	Std
	5 S03	18449	-		5 SFAM01.1	PPB	5/21/2021 8:52	Std
	7.5 S04	27907	-		7.5 SFAM01.1	PPB	5/21/2021 8:54	Std
	10 S05	37748	-		10 SFAM01.1	PPB	5/21/2021 8:57	Std
ICV030	ICV030	14227	3.798 -		SFAM01.1	PPB	5/21/2021 8:59	SMPL
ICB030	ICB030	11	0.0042 -		SFAM01.1	PPB	5/21/2021 9:02	SMPL
CCV041	CCV041	18586	4.9612 -		SFAM01.1	PPB	5/21/2021 9:04	SMPL
CCB041	CCB041	-14	-0.0025 -		SFAM01.1	PPB	5/21/2021 9:06	SMPL
PB136577BL	PBW577	42	0.0124 -		SFAM01.1	PPB	5/21/2021 9:08	SMPL
M2287-01	MHOAM7	27	0.0084 -		SFAM01.1	PPB	5/21/2021 9:11	SMPL
M2287-02	MHOAM7D	25	0.0079 -		SFAM01.1	PPB	5/21/2021 9:13	SMPL
M2287-03	MHOAM7S	3541	0.9462 -		SFAM01.1	PPB	5/21/2021 9:15	SMPL
M2287-04	MHOAM9	10	0.0039 -		SFAM01.1	PPB	5/21/2021 9:17	SMPL
M2287-05	MHOAT4	57	0.0164 -		SFAM01.1	PPB	5/21/2021 9:20	SMPL
M2287-08	MHOAT5	163	0.0447 -		SFAM01.1	PPB	5/21/2021 9:22	SMPL
M2287-09	MHOAT6	51	0.0148 -		SFAM01.1	PPB	5/21/2021 9:24	SMPL
M2287-10	MHOAT7	59	0.017 -		SFAM01.1	PPB	5/21/2021 9:27	SMPL
M2287-11	MHOAT8	46	0.0135 -		SFAM01.1	PPB	5/21/2021 9:29	SMPL
M2287-12	MHOAT9	67	0.0191 -		SFAM01.1	PPB	5/21/2021 9:31	SMPL
M2287-13	MHOAU0	40	0.0119 -		SFAM01.1	PPB	5/21/2021 9:33	SMPL
M2287-14	MHOAU1	32	0.0098 -		SFAM01.1	PPB	5/21/2021 9:36	SMPL
M2287-15	MHOAU2	57	0.0164 -		SFAM01.1	PPB	5/21/2021 9:38	SMPL
M2287-16	MHOAX0	60	0.0172 -		SFAM01.1	PPB	5/21/2021 9:40	SMPL
M2287-17	MHOAX3	147	0.0405 -		SFAM01.1	PPB	5/21/2021 9:42	SMPL
PB136578BL	PBW578	127	0.0351 -		SFAM01.1	PPB	5/21/2021 9:45	SMPL
M2288-01	MHOAQ2	93	0.0261 -		SFAM01.1	PPB	5/21/2021 9:47	SMPL
M2288-02	MHOAM8	57	0.0164 -		SFAM01.1	PPB	5/21/2021 9:49	SMPL
M2288-03	MHOAM8D	66	0.0189 -		SFAM01.1	PPB	5/21/2021 9:51	SMPL
CCV042	CCV042	18382	4.9068 -		SFAM01.1	PPB	5/21/2021 9:54	SMPL
CCB042	CCB042	14	0.005 -		SFAM01.1	PPB	5/21/2021 9:56	SMPL
M2288-04	MHOAM8S	4227	1.1293 -		SFAM01.1	PPB	5/21/2021 9:58	SMPL
M2288-05	MHOAR9	41	0.0122 -		SFAM01.1	PPB	5/21/2021 10:00	SMPL
M2288-06	MHOAS0	65	0.0186 -		SFAM01.1	PPB	5/21/2021 10:03	SMPL
M2288-07	MHOAS1	94	0.0263 -		SFAM01.1	PPB	5/21/2021 10:05	SMPL
M2288-08	MHOAS2	101	0.0282 -		SFAM01.1	PPB	5/21/2021 10:07	SMPL
M2288-09	MHOAS3	304	0.0824 -		SFAM01.1	PPB	5/21/2021 10:09	SMPL
M2288-10	MHOAS4	139	0.0383 -		SFAM01.1	PPB	5/21/2021 10:12	SMPL
M2288-11	MHOAS5	92	0.0258 -		SFAM01.1	PPB	5/21/2021 10:14	SMPL
M2288-12	MHOAS6	71	0.0202 -		SFAM01.1	PPB	5/21/2021 10:16	SMPL
M2288-13	MHOAS7	64	0.0183 -		SFAM01.1	PPB	5/21/2021 10:19	SMPL
M2288-14	MHOAS8	51	0.0148 -		SFAM01.1	PPB	5/21/2021 10:21	SMPL
M2288-15	MHOAS9	49	0.0143 -		SFAM01.1	PPB	5/21/2021 10:23	SMPL
M2288-16	MHOAT0	63	0.018 -		SFAM01.1	PPB	5/21/2021 10:25	SMPL
M2288-17	MHOAT1	66	0.0189 -		SFAM01.1	PPB	5/21/2021 10:28	SMPL

LB114578

INSTRUMENT ID : CV1

M2288-18	MH0AT2	111	0.0309 -	SFAM01.1	PPB	5/21/2021 10:30 SMPL
M2288-19	MH0AW9	88	0.0247 -	SFAM01.1	PPB	5/21/2021 10:32 SMPL
M2288-20	MH0AX1	76	0.0215 -	SFAM01.1	PPB	5/21/2021 10:34 SMPL
CCV043	CCV043	18056	4.8198 -	SFAM01.1	PPB	5/21/2021 10:37 SMPL
CCB043	CCB043	-1	0.001 -	SFAM01.1	PPB	5/21/2021 10:39 SMPL