


LB114521

SFAM01.1

INSTRUMENT ID: CV1

 μ Abs.:

41045



Linear

A= 0.0000e+000

B= 2.4226e-004 *slope*C= 4.7226e-002 *y-intercept*

Rho= 0.9997168

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.057	0.057	40	0.000	40					=
0.05	0.050					204					16
0.20	0.200	0.232	0.032	761	0.0 %	761					-7
2.50	2.500	2.322	-0.178	9391	0.0 %	9391					2
5.00	5.000	5.102	0.102	20864	0.0 %	20864					0
7.50	7.500	7.497	-0.003	30751	0.0 %	30751					0
10.0	10.000	9.991	-0.009	41045	0.0 %	41045					

LB114581
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0	40	-		0 SFAM01.1	PPB	5/21/2021 12:39	Std
	0.2 S01	761	-		0.2 SFAM01.1	PPB	5/21/2021 12:41	Std
	2.5 S02	9391	-		2.5 SFAM01.1	PPB	5/21/2021 12:44	Std
	5 S03	20864	-		5 SFAM01.1	PPB	5/21/2021 12:48	Std
	7.5 S04	30751	-		7.5 SFAM01.1	PPB	5/21/2021 12:50	Std
	10 S05	41045	-		10 SFAM01.1	PPB	5/21/2021 12:53	Std
ICV031	ICV031	15948	3.9107 -		SFAM01.1	PPB	5/21/2021 12:55	SMPL
ICB031	ICB031	-77	0.0286 -		SFAM01.1	PPB	5/21/2021 12:58	SMPL
CCV044	CCV044	20860	5.1007 -		SFAM01.1	PPB	5/21/2021 13:00	SMPL
CCB044	CCB044	-73	0.0295 -		SFAM01.1	PPB	5/21/2021 13:02	SMPL
PB136589BL	PBW589	9	0.0494 -		SFAM01.1	PPB	5/21/2021 13:04	SMPL
M2328-01	MHOAQ9	5	0.0484 -		SFAM01.1	PPB	5/21/2021 13:07	SMPL
M2328-02	MHOAR0	14	0.0506 -		SFAM01.1	PPB	5/21/2021 13:09	SMPL
M2328-03	MHOAR1	37	0.0562 -		SFAM01.1	PPB	5/21/2021 13:11	SMPL
M2328-04	MHOAR2	20	0.0521 -		SFAM01.1	PPB	5/21/2021 13:13	SMPL
M2328-05	MHOAR3	18	0.0516 -		SFAM01.1	PPB	5/21/2021 13:16	SMPL
M2328-06	MHOAR3D	18	0.0516 -		SFAM01.1	PPB	5/21/2021 13:18	SMPL
M2328-07	MHOAR3S	4439	1.1226 -		SFAM01.1	PPB	5/21/2021 13:20	SMPL
M2328-08	MHOAR4	-14	0.0438 -		SFAM01.1	PPB	5/21/2021 13:22	SMPL
M2328-09	MHOAR5	34	0.0555 -		SFAM01.1	PPB	5/21/2021 13:25	SMPL
M2328-10	MHOAR6	18	0.0516 -		SFAM01.1	PPB	5/21/2021 13:27	SMPL
M2328-11	MHOAR7	18	0.0516 -		SFAM01.1	PPB	5/21/2021 13:29	SMPL
M2328-12	MHOAR8	17	0.0513 -		SFAM01.1	PPB	5/21/2021 13:32	SMPL
M2328-14	MHOAX4	31	0.0547 -		SFAM01.1	PPB	5/21/2021 13:34	SMPL
M2328-15	MHOAX5	13	0.0504 -		SFAM01.1	PPB	5/21/2021 13:36	SMPL
M2328-16	MHOAX6	14	0.0506 -		SFAM01.1	PPB	5/21/2021 13:38	SMPL
M2328-17	MHOAX7	16	0.0511 -		SFAM01.1	PPB	5/21/2021 13:41	SMPL
M2328-18	MHOAX8	33	0.0552 -		SFAM01.1	PPB	5/21/2021 13:43	SMPL
M2328-19	MHOAX9	57	0.061 -		SFAM01.1	PPB	5/21/2021 13:45	SMPL
M2328-20	MHOAY0	26	0.0535 -		SFAM01.1	PPB	5/21/2021 13:47	SMPL
CCV045	CCV045	20589	5.0351 -		SFAM01.1	PPB	5/21/2021 13:50	SMPL
CCB045	CCB045	-91	0.0252 -		SFAM01.1	PPB	5/21/2021 13:52	SMPL
M2328-21	MHOAY1	9	0.0494 -		SFAM01.1	PPB	5/21/2021 13:54	SMPL
PB136590BL	PBW590	21	0.0523 -		SFAM01.1	PPB	5/21/2021 13:56	SMPL
M2331-01	MHOAU8	397	0.1434 -		SFAM01.1	PPB	5/21/2021 13:59	SMPL
M2331-02	MHOAU8D	355	0.1332 -		SFAM01.1	PPB	5/21/2021 14:01	SMPL
M2331-03	MHOAU8S	4391	1.111 -		SFAM01.1	PPB	5/21/2021 14:03	SMPL
PB136591BL	PBW591	-2	0.0467 -		SFAM01.1	PPB	5/21/2021 14:05	SMPL
M2332-01	MHOAT3	7	0.0489 -		SFAM01.1	PPB	5/21/2021 14:08	SMPL
M2332-02	MHOAT3D	20	0.0521 -		SFAM01.1	PPB	5/21/2021 14:10	SMPL
M2332-03	MHOAT3S	4497	1.1367 -		SFAM01.1	PPB	5/21/2021 14:12	SMPL
CCV046	CCV046	20378	4.9839 -		SFAM01.1	PPB	5/21/2021 14:14	SMPL
CCB046	CCB046	-94	0.0245 -		SFAM01.1	PPB	5/21/2021 14:17	SMPL