

LB114588


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

49155



A= 0.0000e+000

B= 2.0297e-004

C= -3.9860e-002

Rho= 0.9999357

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.032	-0.032	40	0.000	40					-
0.2	0.200	0.175	-0.025	1059	0.0 %	1059					-B
2.5	2.500	2.530	0.030	12661	0.0 %	12661					-
5.0	5.000	5.049	0.049	25074	0.0 %	25074					-
7.5	7.500	7.540	0.040	37344	0.0 %	37344					-
10.0	10.000	9.937	-0.063	49155	0.0 %	49155					-

LB114588 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 May 2021 15:54:55

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	40	-	0.00007471B	PPB	21 May 2021 15:55:03	S	Std
0.2 - 1	50.2	1059	-	0.20007471B	PPB	21 May 2021 15:57:20	S	Std
2.5 - 1	52.5	12661	-	2.50007471B	PPB	21 May 2021 15:59:36	S	Std
5.0 - 1	55	25074	-	5.00007471B	PPB	21 May 2021 16:01:53	S	Std
7.5 - 1	57.5	37344	-	7.50007471B	PPB	21 May 2021 16:04:10	S	Std
10.0 - 1	510	49155	-	10.00007471B	PPB	21 May 2021 16:06:26	S	Std
ICV37 - 1	ICV37	19731	3.9650	-7471B	PPB	21 May 2021 16:11:49	U	SMPL
ICB37 - 1	ICB37	-89	-0.0579	-7471B	PPB	21 May 2021 16:14:04	U	SMPL
CCV49 - 1	CCV49	25351	5.1057	-7471B	PPB	21 May 2021 16:16:20	U	SMPL
CCB49 - 1	CCB49	-109	-0.0620	-7471B	PPB	21 May 2021 16:18:35	U	SMPL
CRA - 1	CRA	1047	0.1727	-7471B	PPB	21 May 2021 16:20:52	U	SMPL
HighStd - 1	HighStd	49418	9.9906	-7471B	PPB	21 May 2021 16:23:07	U	SMPL
ChkStd - 1	ChkStd	34019	6.8651	-7471B	PPB	21 May 2021 16:25:22	U	SMPL
PB136599BL - 1	PBS	-165	-0.0734	-7471B	PPB	21 May 2021 16:27:41	U	SMPL
PB136599BS - 1	LCSS	20271	4.0746	-7471B	PPB	21 May 2021 16:29:58	U	SMPL
M2479-01 - 1	SUBBASE-6-1-D	288	0.0186	-7471B	PPB	21 May 2021 16:32:15	U	SMPL
M2479-02 - 1	MIXED-SUBBASE-GROUND-1-15-D	858	0.1343	-7471B	PPB	21 May 2021 16:34:32	U	SMPL
M2479-03 - 1	GROUND-MATERIAL-1-17-D	375	0.0363	-7471B	PPB	21 May 2021 16:36:48	U	SMPL
M2479-04 - 1	MIXED-SUBBASE-GROUND-4-10-D	5867	1.1510	-7471B	PPB	21 May 2021 16:39:05	U	SMPL
CCV50 - 1	CCV50	25294	5.0941	-7471B	PPB	21 May 2021 16:41:20	U	SMPL
CCB50 - 1	CCB50	-124	-0.0650	-7471B	PPB	21 May 2021 16:43:35	U	SMPL
M2482-01 - 1	BU-02-052021	1282	0.2204	-7471B	PPB	21 May 2021 16:45:51	U	SMPL
M2482-01DUP - 1	BU-02-052021DUP	1278	0.2195	-7471B	PPB	21 May 2021 16:48:06	U	SMPL
M2482-01MS - 1	BU-02-052021MS	19731	3.9650	-7471B	PPB	21 May 2021 16:50:22	U	SMPL
M2482-01MSD - 1	BU-02-052021MSD	19732	3.9652	-7471B	PPB	21 May 2021 16:52:37	U	SMPL
M2486-01 - 1	OR-03-052121	3690	0.7091	-7471B	PPB	21 May 2021 16:54:52	U	SMPL
M2490-01 - 1	72-11986	4475	0.8684	-7471B	PPB	21 May 2021 16:57:09	U	SMPL
M2482-01LX5 - 1		239	0.0087	-7471B	PPB	21 May 2021 16:59:25	U	SMPL
M2482-01A - 1		21508	4.3257	-7471B	PPB	21 May 2021 17:01:41	U	SMPL
CCV51 - 1	CCV51	25192	5.0734	-7471B	PPB	21 May 2021 17:03:58	U	SMPL
CCB51 - 1	CCB51	-98	-0.0598	-7471B	PPB	21 May 2021 17:06:15	U	SMPL