



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114590

Slope : 99.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	23.1	4.01	05/21/2021	15:25
2	CAL2	1	Water	NA	NA	23.2	7.01	05/21/2021	15:26
3	CAL3	1	Water	NA	NA	23.2	10.03	05/21/2021	15:27
4	ICV	1	Water	NA	NA	23.1	6.98	05/21/2021	15:29
5	CCV1	1	Water	NA	NA	23.2	2.01	05/21/2021	15:30
6	M2489-02	1	Solid	20.02	20	25.6	10.90	05/21/2021	15:33
7	M2489-02DUP	1	Solid	20.03	20	25.7	10.91	05/21/2021	15:34
8	CCV2	1	Water	NA	NA	23.3	12.03	05/21/2021	15:36

13116390

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY M2489 WorkList ID : 149146 Department : Wet-Chemistry Date : 05-21-2021 15:15:07

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/21/2021	Solid	M2489-02	Corrosivity	Cool 4 deg C	PSEG03	F52	359	05/21/2021	9045D

Date/Time 05/21/21 15:20
Raw Sample Received by: JO (Caf-chem)
Raw Sample Relinquished by: JT (S)

Date/Time 05/21/21 16:23
Raw Sample Received by: JT (S)
Raw Sample Relinquished by: JO Caf-chem