



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB114591

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP90886
COD calibration std. 100 ppm	WP90884
COD calibration std. 10 ppm	WP90882
COD calibration std. 150 ppm	WP90885
COD calibration std. 50 ppm	WP90883
COD calibration std. 0 ppm	WP90881
COD Digestion Vials Low Level 0-150Mg/L	W2782
COD ICV-LCS std, 50ppm	WP91215
COD CCV std, 50ppm	WP91214

Temp In (C): <u>148</u>	Date In: <u>05/21/2021</u>	Time In: <u>09:32</u>
Temp Out (C): <u>150</u>	Date Out: <u>05/21/2021</u>	Time Out: <u>11:32</u>

Intercept: 0.2708

Slope: 1.015

Regression: 1

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	-0.267		05/08/2021	12:05
2	CAL2	10	1	Water	11.000	10.571	-5.7	05/08/2021	12:05
3	CAL3	50	1	Water	51.000	49.98	0	05/08/2021	12:06
4	CAL4	100	1	Water	101.000	99.241	0.8	05/08/2021	12:06
5	CAL5	150	1	Water	153.000	150.472	-0.3	05/08/2021	12:07



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Alexander

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB114591

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	52.000	50.965	05/08/2021	12:08
2	ICB				1	Water	0.000	-0.267	05/08/2021	12:08
3	CCV1	50			1	Water	51.000	49.980	05/21/2021	15:20
4	CCB1				1	Water	0.000	-0.267	05/21/2021	15:20
5	LB114591BL				1	Water	1.000	0.718	05/21/2021	15:21
6	LB114591BS	50			1	Water	52.000	50.965	05/21/2021	15:21
7	M2483-01				1	Water	22.000	21.408	05/21/2021	15:22
8	M2483-01DUP				1	Water	24.000	23.379	05/21/2021	15:23
9	M2483-01MS	50			1	Water	67.000	65.743	05/21/2021	15:23
10	M2483-01MSD	50			1	Water	69.000	67.713	05/21/2021	15:24
11	M2483-02				10	Water	123.000	120.915	05/21/2021	15:24
12	CCV2	50			1	Water	52.000	50.965	05/21/2021	15:25
13	CCB2				1	Water	1.000	0.718	05/21/2021	15:25

LB114591
WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD-52121 WorkList ID : 149125 Department : Wet-Chemistry Date : 05-21-2021 08:52:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/03/2021	Water	M2483-01	COD	Conc H2SO4 1c	TRIS02	F11	DSN-001-SEWER-OUTLET	05/20/2021	SM5220 D
06/03/2021	Water	M2483-02	COD	Conc H2SO4 1c	TRIS02	F11	DSN-002-SEWER-OUTLET	05/20/2021	SM5220 D

Date/Time 05/21/2021 9:05
 Raw Sample Received by: ARLWet-chem
 Raw Sample Relinquished by: DR Cef Chem

Date/Time 05/21/2021 9:40
 Raw Sample Received by: ARLWet-chem
 Raw Sample Relinquished by: ARLWet-chem