

LB 114 602


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

42990



A= 0.0000e+000

B= 2.3323e-004 slope

C= -6.9398e-002 y-intercept

Rho= 0.9999515

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.001	0.001	302	0.000	302					
0.05	0.050					204					
0.20	0.200	0.187	-0.013	1099	0.0 %	1099					
2.50	2.500	2.466	-0.034	10869	0.0 %	10869					
5.00	5.000	5.064	0.064	22011	0.0 %	22011					
7.50	7.500	7.525	0.025	32560	0.0 %	32560					
10.0	10.000	9.957	-0.043	42990	0.0 %	42990					

100
117
117
117
117
117

LB114602
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	302	-		0 SFAM01.1	PPB	5/24/2021 13:42	Std
	0.2 S01	1099	-		0.2 SFAM01.1	PPB	5/24/2021 13:45	Std
	2.5 S02	10869	-		2.5 SFAM01.1	PPB	5/24/2021 13:47	Std
	5 S03	22011	-		5 SFAM01.1	PPB	5/24/2021 13:49	Std
	7.5 S04	32560	-		7.5 SFAM01.1	PPB	5/24/2021 13:51	Std
	10 S05	42990	-		10 SFAM01.1	PPB	5/24/2021 13:54	Std
ICV032	ICV032	16714	3.8289 -		SFAM01.1	PPB	5/24/2021 13:56	SMPL
ICB032	ICB032	177	-0.0281 -		SFAM01.1	PPB	5/24/2021 13:59	SMPL
CCV052	CCV052	21699	4.9916 -		SFAM01.1	PPB	5/24/2021 14:01	SMPL
CCB052	CCB052	178	-0.0279 -		SFAM01.1	PPB	5/24/2021 14:03	SMPL
PB136610BL	PBW610	294	-0.0008 -		SFAM01.1	PPB	5/24/2021 14:05	SMPL
M2318-01	MA0AW0	315	0.0041 -		SFAM01.1	PPB	5/24/2021 14:08	SMPL
M2318-02	MA0AZ8	314	0.0038 -		SFAM01.1	PPB	5/24/2021 14:10	SMPL
M2318-03	MA0AL8	324	0.0062 -		SFAM01.1	PPB	5/24/2021 14:12	SMPL
M2318-04	MA0AN2	321	0.0055 -		SFAM01.1	PPB	5/24/2021 14:14	SMPL
M2318-05	MA0AY6	292	-0.0013 -		SFAM01.1	PPB	5/24/2021 14:17	SMPL
M2318-06	MA0B00	229	-0.016 -		SFAM01.1	PPB	5/24/2021 14:19	SMPL
M2318-07	MA0AW2	273	-0.0057 -		SFAM01.1	PPB	5/24/2021 14:21	SMPL
M2318-08	MA0AW4	279	-0.0043 -		SFAM01.1	PPB	5/24/2021 14:24	SMPL
M2318-09	MA0AW6	215	-0.0193 -		SFAM01.1	PPB	5/24/2021 14:26	SMPL
M2318-10	MA0B02	573	0.0642 -		SFAM01.1	PPB	5/24/2021 14:28	SMPL
M2318-11	MA0AM0	306	0.002 -		SFAM01.1	PPB	5/24/2021 14:30	SMPL
M2318-12	MA0AP0	263	-0.0081 -		SFAM01.1	PPB	5/24/2021 14:33	SMPL
M2318-13	MA0AP4	285	-0.0029 -		SFAM01.1	PPB	5/24/2021 14:35	SMPL
M2318-14	MA0AQ0	320	0.0052 -		SFAM01.1	PPB	5/24/2021 14:37	SMPL
M2318-15	MA0AQ2	215	-0.0193 -		SFAM01.1	PPB	5/24/2021 14:39	SMPL
M2318-16	MA0AR0	156	-0.033 -		SFAM01.1	PPB	5/24/2021 14:42	SMPL
M2318-17	MA0AM2	186	-0.026 -		SFAM01.1	PPB	5/24/2021 14:44	SMPL
M2318-18	MA0AP6	427	0.0302 -		SFAM01.1	PPB	5/24/2021 14:46	SMPL
M2318-19	MA0AZ4	363	0.0153 -		SFAM01.1	PPB	5/24/2021 14:48	SMPL
CCV053	CCV053	21472	4.9386 -		SFAM01.1	PPB	5/24/2021 14:51	SMPL
CCB053	CCB053	185	-0.0263 -		SFAM01.1	PPB	5/24/2021 14:53	SMPL
M2318-20	MA0AZ6	296	-0.0004 -		SFAM01.1	PPB	5/24/2021 14:55	SMPL
M2318-21	MA0AZ6D	331	0.0078 -		SFAM01.1	PPB	5/24/2021 14:57	SMPL
M2318-22	MA0AZ6S	5210	1.1458 -		SFAM01.1	PPB	5/24/2021 15:00	SMPL
PB136611BL	PBW611	284	-0.0032 -		SFAM01.1	PPB	5/24/2021 15:02	SMPL
M2319-01	MA0AW1	364	0.0155 -		SFAM01.1	PPB	5/24/2021 15:04	SMPL
M2319-02	MA0AZ9	335	0.0087 -		SFAM01.1	PPB	5/24/2021 15:06	SMPL
M2319-03	MA0AL9	274	-0.0055 -		SFAM01.1	PPB	5/24/2021 15:09	SMPL
M2319-04	MA0AN3	308	0.0024 -		SFAM01.1	PPB	5/24/2021 15:11	SMPL
M2319-05	MA0AY7	417	0.0279 -		SFAM01.1	PPB	5/24/2021 15:13	SMPL
M2319-06	MA0B01	299	0.0003 -		SFAM01.1	PPB	5/24/2021 15:16	SMPL
M2319-07	MA0AW3	299	0.0003 -		SFAM01.1	PPB	5/24/2021 15:18	SMPL
M2319-08	MA0AW5	308	0.0024 -		SFAM01.1	PPB	5/24/2021 15:20	SMPL
M2319-09	MA0AW7	284	-0.0032 -		SFAM01.1	PPB	5/24/2021 15:22	SMPL
M2319-10	MA0B03	291	-0.0015 -		SFAM01.1	PPB	5/24/2021 15:25	SMPL

LB114602

INSTRUMENT ID : CV1

M2319-11	MA0AM1	319	0.005 -	SFAM01.1	PPB	5/24/2021 15:27 SMPL
M2319-12	MA0AP1	313	0.0036 -	SFAM01.1	PPB	5/24/2021 15:29 SMPL
M2319-13	MA0AP5	340	0.0099 -	SFAM01.1	PPB	5/24/2021 15:31 SMPL
M2319-14	MA0AQ1	308	0.0024 -	SFAM01.1	PPB	5/24/2021 15:34 SMPL
M2319-15	MA0AQ3	316	0.0043 -	SFAM01.1	PPB	5/24/2021 15:36 SMPL
M2319-16	MA0AR1	274	-0.0055 -	SFAM01.1	PPB	5/24/2021 15:38 SMPL
CCV054	CCV054	21255	4.888 -	SFAM01.1	PPB	5/24/2021 15:40 SMPL
CCB054	CCB054	173	-0.029 -	SFAM01.1	PPB	5/24/2021 15:43 SMPL
M2319-17	MA0AM3	271	-0.0062 -	SFAM01.1	PPB	5/24/2021 15:45 SMPL
M2319-18	MA0AP7	298	0.0001 -	SFAM01.1	PPB	5/24/2021 15:47 SMPL
M2319-19	MA0AZ5	296	-0.0004 -	SFAM01.1	PPB	5/24/2021 15:50 SMPL
M2319-20	MA0AZ7	294	-0.0008 -	SFAM01.1	PPB	5/24/2021 15:52 SMPL
M2319-21	MA0AZ7D	290	-0.0018 -	SFAM01.1	PPB	5/24/2021 15:54 SMPL
M2319-22	MA0AZ7S	4472	0.9736 -	SFAM01.1	PPB	5/24/2021 15:56 SMPL
CCV055	CCV055	21510	4.9475 -	SFAM01.1	PPB	5/24/2021 15:59 SMPL
CCB055	CCB055	173	-0.029 -	SFAM01.1	PPB	5/24/2021 16:01 SMPL