



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 5/25/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 16:00
In Date: 05/24/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:24
Out Date: 05/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114603

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
M2422-07	K328-12B	1	1.15	9.82	8.1	80.2	
M2422-10	K328-13B	2	1.16	9.79	8.39	83.8	
M2422-13	K328-14B	3	1.12	9.88	8.45	83.7	
M2422-16	K328-15B	4	1.13	9.54	8.21	84.2	
M2422-16D	K328-15BDUP	5	1.15	9.56	8.23	84.2	
M2498-03	BP-VPB-RW8-GW-958-960	6	1.14	9.96	2.85	19.4	SLUDGE SAMPLE
M2499-01	C1	7	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2499-02	C2	8	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2499-03	C3	9	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2499-04	C4	10	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2499-05	C5	11	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2499-06	C6	12	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2500-01	NB-08-052421	15	1.18	9.61	9.37	97.2	
M2500-02	NB-08-052421-EPH-2	16	1.16	9.65	9.38	96.8	
M2502-01	SA-01-052421	17	1.13	9.58	9.32	96.9	
M2502-02	SA-01-052421-EPH-2	18	1.17	9.69	9.39	96.5	
M2503-01	OK-01-052421	13	1.15	9.54	9.05	94.2	
M2503-02	OK-01-052421-EPH-2	14	1.11	9.88	9.14	91.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$