



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB114605

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP91271
Turbidity Calibration std, 0NTU	WP91270
Turbidity Calibration std, 20NTU	WP91266
Turbidity Calibration std, 40NTU	WP91265
Turbidity Calibration std, 1NTU	WP91269
Turbidity Calibration std, 5NTU	WP91268
Turbidity LOQ std, 1.0NTU	WP91272
Turbidity Calibration - CCV std, 10 NTU	WP91267
10 NTU Standard 500 ml	W2752

Intercept: 0.0947

Slope: 0.9722

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.205	0.11		05/24/2021	09:40
2	CAL2	1	1	1.130	1.07	-6.5	05/24/2021	09:42
3	CAL3	5	1	5.050	5.10	-1.9	05/24/2021	09:44
4	CAL4	10	1	9.770	9.95	0.5	05/24/2021	09:46
5	CAL5	20	1	19.100	19.55	2.3	05/24/2021	09:48
6	CAL6	40	1	39.200	40.22	-0.6	05/24/2021	09:50



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB114605

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.291	05/24/2021	09:53
2	ICB		1	0.254	0.164	05/24/2021	09:55
3	CCV1	10	1	9.660	9.839	05/24/2021	09:57
4	CCB1		1	0.269	0.179	05/24/2021	09:59
5	LB114605BL		1	0.277	0.188	05/24/2021	10:02
6	M2262-07		1	0.688	0.610	05/24/2021	10:04
7	M2262-08		1	1.200	1.137	05/24/2021	10:06
8	CCV2	10	1	9.810	9.993	05/24/2021	10:08
9	CCB2		1	0.282	0.193	05/24/2021	10:10