

LB114611


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51353



A= 0.0000e+000

B= 1.9548e-004

C= 2.4353e-002

Rho= 0.9998946

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%OD
0.0	0.000	0.078	0.078	273	0.000	273					1
0.2	0.200	0.190	-0.010	849	0.0 %	849					-5
2.5	2.500	2.439	-0.061	12352	0.0 %	12352					-2
5.0	5.000	4.987	-0.013	25387	0.0 %	25387					0
7.5	7.500	7.443	-0.057	37951	0.0 %	37951					-1
10.0	10.000	10.063	0.063	51353	0.0 %	51353					1

LB114611 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 May 2021 16:10:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	273	-	0.0000	7471B	PPB	24 May 2021 16:11:07	S	Std
0.2 - 1	50.2	849	-	0.2000	7471B	PPB	24 May 2021 16:13:24	S	Std
2.5 - 1	52.5	12352	-	2.5000	7471B	PPB	24 May 2021 16:15:41	S	Std
5.0 - 1	55	25387	-	5.0000	7471B	PPB	24 May 2021 16:17:57	S	Std
7.5 - 1	57.5	37951	-	7.5000	7471B	PPB	24 May 2021 16:20:13	S	Std
10.0 - 1	57.0	51353	-	10.0000	7471B	PPB	24 May 2021 16:22:30	S	Std
ICV38 - 1	ICV38	20009	3.9357	-	7471B	PPB	24 May 2021 16:25:54	U	SMPL
ICB38 - 1	ICB38	-24	0.0197	-	7471B	PPB	24 May 2021 16:28:10	U	SMPL
CCV56 - 1	CCV56	25324	4.9747	-	7471B	PPB	24 May 2021 16:30:26	U	SMPL
CCB56 - 1	CCB56	-29	0.0187	-	7471B	PPB	24 May 2021 16:32:41	U	SMPL
CRA - 1	CRA	977	0.2153	-	7471B	PPB	24 May 2021 16:34:58	U	SMPL
HighStd - 1	HighStd	51402	10.0725	-	7471B	PPB	24 May 2021 16:37:12	U	SMPL
ChkStd - 1	ChkStd	36110	7.0832	-	7471B	PPB	24 May 2021 16:39:27	U	SMPL
PB136624BL - 1	PBS	-74	0.0099	-	7471B	PPB	24 May 2021 16:41:45	U	SMPL
PB136624BS - 1	LCSS	18658	3.6716	-	7471B	PPB	24 May 2021 16:44:02	U	SMPL
M2485-01 - 1	CL-01-052121	2718	0.5557	-	7471B	PPB	24 May 2021 16:46:18	U	SMPL
M2485-01DUP - 1	CL-01-052121DUP	2753	0.5625	-	7471B	PPB	24 May 2021 16:48:35	U	SMPL
M2485-01MS - 1	CL-01-052121MS	23854	4.6874	-	7471B	PPB	24 May 2021 16:53:32	U	SMPL
M2485-01MSD - 1	CL-01-052121MSD	22457	4.4143	-	7471B	PPB	24 May 2021 16:55:48	U	SMPL
M2489-02 - 1	359	3950	0.7965	-	7471B	PPB	24 May 2021 16:58:04	U	SMPL
CCV57 - 1	CCV57	25336	4.9771	-	7471B	PPB	24 May 2021 17:00:20	U	SMPL
CCB57 - 1	CCB57	-33	0.0179	-	7471B	PPB	24 May 2021 17:02:35	U	SMPL
M2492-01 - 1	OL-STA-80082-0-4	7720	1.5335	-	7471B	PPB	24 May 2021 17:04:51	U	SMPL
M2492-02 - 1	OL-STA-80082-4-10	23846	4.6858	-	7471B	PPB	24 May 2021 17:07:06	U	SMPL
M2492-03 - 1	ST-51-0-4	10919	2.1588	-	7471B	PPB	24 May 2021 17:09:21	U	SMPL
M2492-04 - 1	OL-STA-80238-4-10	55654	10.9037	-	7471B	PPB	24 May 2021 17:11:38	U	SMPL
M2492-05 - 1	OL-STA-80083-4-10	19941	3.9224	-	7471B	PPB	24 May 2021 17:13:54	U	SMPL
M2492-06 - 1	S-94-0-6	847	0.1899	-	7471B	PPB	24 May 2021 17:16:12	U	SMPL
M2492-07 - 1	S-93-0-6	5728	1.1441	-	7471B	PPB	24 May 2021 17:18:28	U	SMPL
M2492-08 - 1	S-67-0-6	4551	0.9140	-	7471B	PPB	24 May 2021 17:20:46	U	SMPL
M2492-09 - 1	S-364-0-6	2637	0.5398	-	7471B	PPB	24 May 2021 17:23:03	U	SMPL
M2492-10 - 1	S-361-0-6	6595	1.3136	-	7471B	PPB	24 May 2021 17:25:19	U	SMPL
CCV58 - 1	CCV58	25188	4.9481	-	7471B	PPB	24 May 2021 17:27:34	U	SMPL
CCB58 - 1	CCB58	-9	0.0226	-	7471B	PPB	24 May 2021 17:29:49	U	SMPL
M2485-01LX5 - 1		595	0.1407	-	7471B	PPB	24 May 2021 17:32:06	U	SMPL
M2485-01A - 1		22772	4.4759	-	7471B	PPB	24 May 2021 17:34:21	U	SMPL
M2492-04DLX2 - 1		28078	5.5131	-	7471B	PPB	24 May 2021 17:36:37	U	SMPL
CCV59 - 1	CCV59	25259	4.9620	-	7471B	PPB	24 May 2021 17:38:53	U	SMPL
CCB59 - 1	CCB59	-3	0.0238	-	7471B	PPB	24 May 2021 17:41:10	U	SMPL