



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB114618

Reagent/Standard	Lot/Log #
TURBIDITY LOD STD, 0.5NTU	WP91309
Turbidity Calibration std, 0NTU	WP91308
Turbidity Calibration std, 20NTU	WP91304
Turbidity Calibration std, 40NTU	WP91303
Turbidity Calibration std, 1NTU	WP91307
Turbidity Calibration std, 5NTU	WP91306
Turbidity Calibration - CCV std, 10 NTU	WP91305
10 NTU Standard 500 ml	W2752

Intercept: 0.1201

Slope: 0.9606

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.214	0.10		05/25/2021	09:48
2	CAL2	1	1	1.150	1.07	-7.2	05/25/2021	09:50
3	CAL3	5	1	5.090	5.17	-3.5	05/25/2021	09:52
4	CAL4	10	1	9.670	9.94	0.6	05/25/2021	09:54
5	CAL5	20	1	18.800	19.45	2.8	05/25/2021	09:56
6	CAL6	40	1	38.800	40.27	-0.7	05/25/2021	09:58



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB114618

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.200	10.493	05/25/2021	10:00
2	ICB		1	0.235	0.120	05/25/2021	10:02
3	CCV1	10	1	9.850	10.129	05/25/2021	10:05
4	CCB1		1	0.249	0.134	05/25/2021	10:08
5	LB114618BL		1	0.265	0.151	05/25/2021	10:10
6	M2262-12		1	0.628	0.529	05/25/2021	10:12
7	CCV2	10	1	9.356	9.615	05/25/2021	10:14
8	CCB2		1	0.254	0.139	05/25/2021	10:17