


LB114621

SFAM01.1

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.4508e-004 slope

C= -6.9737e-002 y-intercept

Rho= 0.9999659

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.011	-0.011	238	0.000	238					
0.05	0.050					204					
0.20	0.200	0.186	-0.014	1043	0.0 %	1043					
2.50	2.500	2.498	-0.002	10479	0.0 %	10479					
5.00	5.000	5.032	0.032	20818	0.0 %	20818					
7.50	7.500	7.543	0.043	31062	0.0 %	31062					
10.0	10.000	9.952	-0.048	40890	0.0 %	40890					

0.00
= 7
0
1
1
0

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	238	-		0 SFAM01.1	PPB	5/25/2021 10:42	Std
	0.2 S01	1043	-		0.2 SFAM01.1	PPB	5/25/2021 10:44	Std
	2.5 S02	10479	-		2.5 SFAM01.1	PPB	5/25/2021 10:47	Std
	5 S03	20818	-		5 SFAM01.1	PPB	5/25/2021 10:49	Std
	7.5 S04	31062	-		7.5 SFAM01.1	PPB	5/25/2021 10:51	Std
	10 S05	40890	-		10 SFAM01.1	PPB	5/25/2021 10:53	Std
ICV033	ICV033	16003	3.8523 -		SFAM01.1	PPB	5/25/2021 10:56	SMPL
ICB033	ICB033	138	-0.0359 -		SFAM01.1	PPB	5/25/2021 11:01	SMPL
CCV060	CCV060	19650	4.7461 -		SFAM01.1	PPB	5/25/2021 11:04	SMPL
CCB060	CCB060	-71	-0.0871 -		SFAM01.1	PPB	5/25/2021 11:06	SMPL
PB136634BL	PBW634	1	-0.0695 -		SFAM01.1	PPB	5/25/2021 11:08	SMPL
M2416-01	MH0AY8	68	-0.0531 -		SFAM01.1	PPB	5/25/2021 11:10	SMPL
M2416-02	MH0AZ0	52	-0.057 -		SFAM01.1	PPB	5/25/2021 11:13	SMPL
M2416-03	MH0AZ2	47	-0.0582 -		SFAM01.1	PPB	5/25/2021 11:15	SMPL
M2416-04	MH0AZ4	54	-0.0565 -		SFAM01.1	PPB	5/25/2021 11:17	SMPL
M2416-05	MH0AZ6	45	-0.0587 -		SFAM01.1	PPB	5/25/2021 11:19	SMPL
M2416-06	MH0AZ8	30	-0.0624 -		SFAM01.1	PPB	5/25/2021 11:22	SMPL
M2416-07	MH0BA0	51	-0.0572 -		SFAM01.1	PPB	5/25/2021 11:24	SMPL
M2416-08	MH0BA2	53	-0.0567 -		SFAM01.1	PPB	5/25/2021 11:26	SMPL
M2416-09	MH0BA4	75	-0.0514 -		SFAM01.1	PPB	5/25/2021 11:28	SMPL
M2416-10	MH0BA6	64	-0.0541 -		SFAM01.1	PPB	5/25/2021 11:31	SMPL
M2416-11	MH0BA6D	66	-0.0536 -		SFAM01.1	PPB	5/25/2021 11:33	SMPL
M2416-12	MH0BA6S	4130	0.9425 -		SFAM01.1	PPB	5/25/2021 11:35	SMPL
M2416-13	MH0BA8	34	-0.0614 -		SFAM01.1	PPB	5/25/2021 11:37	SMPL
M2416-14	MH0BB0	52	-0.057 -		SFAM01.1	PPB	5/25/2021 11:40	SMPL
M2416-15	MH0BB2	25	-0.0636 -		SFAM01.1	PPB	5/25/2021 11:42	SMPL
M2416-16	MH0BB4	18	-0.0653 -		SFAM01.1	PPB	5/25/2021 11:44	SMPL
M2416-17	MH0BB6	37	-0.0607 -		SFAM01.1	PPB	5/25/2021 11:47	SMPL
M2416-18	MH0BB8	21	-0.0646 -		SFAM01.1	PPB	5/25/2021 11:49	SMPL
PB136635BL	PBW635	34	-0.0614 -		SFAM01.1	PPB	5/25/2021 11:51	SMPL
CCV061	CCV061	20273	4.8988 -		SFAM01.1	PPB	5/25/2021 11:53	SMPL
CCB061	CCB061	-88	-0.0913 -		SFAM01.1	PPB	5/25/2021 11:56	SMPL
M2417-01	MH0AY9	43	-0.0592 -		SFAM01.1	PPB	5/25/2021 11:58	SMPL
M2417-02	MH0AZ1	127	-0.0386 -		SFAM01.1	PPB	5/25/2021 12:00	SMPL
M2417-03	MH0AZ3	137	-0.0362 -		SFAM01.1	PPB	5/25/2021 12:02	SMPL
M2417-04	MH0AZ5	17	-0.0656 -		SFAM01.1	PPB	5/25/2021 12:05	SMPL
M2417-05	MH0AZ7	16	-0.0658 -		SFAM01.1	PPB	5/25/2021 12:07	SMPL
M2417-06	MH0AZ9	32	-0.0619 -		SFAM01.1	PPB	5/25/2021 12:09	SMPL
M2417-07	MH0BA1	24	-0.0639 -		SFAM01.1	PPB	5/25/2021 12:11	SMPL
M2417-08	MH0BA3	11	-0.067 -		SFAM01.1	PPB	5/25/2021 12:14	SMPL
M2417-09	MH0BA5	49	-0.0577 -		SFAM01.1	PPB	5/25/2021 12:16	SMPL
M2417-10	MH0BA7	28	-0.0629 -		SFAM01.1	PPB	5/25/2021 12:18	SMPL
M2417-11	MH0BA7D	46	-0.0585 -		SFAM01.1	PPB	5/25/2021 12:21	SMPL
M2417-12	MH0BA7S	4086	0.9317 -		SFAM01.1	PPB	5/25/2021 12:23	SMPL
M2417-13	MH0BA9	15	-0.0661 -		SFAM01.1	PPB	5/25/2021 12:25	SMPL
M2417-14	MH0BB1	49	-0.0577 -		SFAM01.1	PPB	5/25/2021 12:27	SMPL

LB114621

INSTRUMENT ID : CV1

M2417-15	MH0BB3	50	-0.0575 -	SFAM01.1	PPB	5/25/2021 12:30 SMPL
M2417-16	MH0BB5	68	-0.0531 -	SFAM01.1	PPB	5/25/2021 12:32 SMPL
M2417-17	MH0BB7	70	-0.0526 -	SFAM01.1	PPB	5/25/2021 12:34 SMPL
M2417-18	MH0BB9	58	-0.0555 -	SFAM01.1	PPB	5/25/2021 12:36 SMPL
CCV062	CCV062	20079	4.8513 -	SFAM01.1	PPB	5/25/2021 12:41 SMPL
CCB062	CCB062	-80	-0.0893 -	SFAM01.1	PPB	5/25/2021 12:43 SMPL