

LB 114635


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34942



A= 0.0000e+000

B= 2.8105e-004

C= 3.1129e-003

Rho= 0.9994130

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/00
0.0	0.000	0.009	0.009	21	0.000	21					1
0.2	0.200	0.194	-0.006	679	0.0 %	679					-3
2.5	2.500	2.375	-0.125	8439	0.0 %	8439					-5
5.0	5.000	5.103	0.103	18145	0.0 %	18145					2
7.5	7.500	7.696	0.196	27372	0.0 %	27372					3
10.0	10.000	9.823	-0.177	34942	0.0 %	34942					-2

LB114633 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 25 May 2021 16:15:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	21	-	0.0000	7470A	PPB	25 May 2021 16:16:44	S	Std
0.2 - 1	50.2	679	-	0.2000	7470A	PPB	25 May 2021 16:19:01	S	Std
2.5 - 1	52.5	8439	-	2.5000	7470A	PPB	25 May 2021 16:21:18	S	Std
5.0 - 1	55	18145	-	5.0000	7470A	PPB	25 May 2021 16:23:35	S	Std
7.5 - 1	57.5	27372	-	7.5000	7470A	PPB	25 May 2021 16:25:51	S	Std
10.0 - 1	50	34942	-	10.0000	7470A	PPB	25 May 2021 16:30:41	S	Std
ICV40 - 1	ICV40	13825	3.8886	-	7470A	PPB	25 May 2021 16:33:22	U	SMPL
ICB40 - 1	ICB40	-117	-0.0298	-	7470A	PPB	25 May 2021 16:35:37	U	SMPL
CCV67 - 1	CCV67	17855	5.0212	-	7470A	PPB	25 May 2021 16:37:52	U	SMPL
CCB67 - 1	CCB67	-116	-0.0295	-	7470A	PPB	25 May 2021 16:40:06	U	SMPL
CRA - 1	CRA	643	0.1838	-	7470A	PPB	25 May 2021 16:42:22	U	SMPL
HighStd - 1	HighStd	34124	9.5936	-	7470A	PPB	25 May 2021 16:44:37	U	SMPL
ChkStd - 1	ChkStd	24006	6.7499	-	7470A	PPB	25 May 2021 16:46:55	U	SMPL
PB136640BL - 1	PBW	-153	-0.0399	-	7470A	PPB	25 May 2021 16:49:13	U	SMPL
PB136640BS - 1	LCSW	15576	4.3807	-	7470A	PPB	25 May 2021 16:51:33	U	SMPL
M2489-03 - 1	359	-82	-0.0199	-	7470A	PPB	25 May 2021 16:53:49	U	SMPL
M2489-03DUP - 1	359DUP	8	0.0054	-	7470A	PPB	25 May 2021 16:56:06	U	SMPL
M2489-03MS - 1	359MS	14297	4.0213	-	7470A	PPB	25 May 2021 16:58:23	U	SMPL
M2489-03MSD - 1	359MSD	13949	3.9234	-	7470A	PPB	25 May 2021 17:00:40	U	SMPL
CCV68 - 1	CCV68	17863	5.0235	-	7470A	PPB	25 May 2021 17:02:55	U	SMPL
CCB68 - 1	CCB68	-133	-0.0343	-	7470A	PPB	25 May 2021 17:05:10	U	SMPL
PB136640TB - 1	PB136614TB	14	0.0070	-	7470A	PPB	25 May 2021 17:07:26	U	SMPL
M2489-03LX5 - 1		10	0.0059	-	7470A	PPB	25 May 2021 17:09:41	U	SMPL
CCV69 - 1	CCV69	13871	3.9015	-	7470A	PPB	25 May 2021 17:11:57	U	SMPL
CCB69 - 1	CCB69	17642	4.9614	-	7470A	PPB	25 May 2021 17:14:12	U	SMPL
		-110	-0.0278	-	7470A	PPB	25 May 2021 17:16:28	U	SMPL