



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 6/2/2021

OVENTEMP IN Celsius(°C): 107
Time IN: 17:20
In Date: 05/25/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:25
Out Date: 05/26/2021
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB114634

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
M2512-01	COMP-1-3	15	1.00	2.00	2.00	100.0	OIL SAMPLE
M2512-04	COMP-4-6	16	1.00	2.00	2.00	100.0	OIL SAMPLE
M2515-01	COMPOSITE-CONCRETE	1	1.00	2.00	2.00	100.0	CONCRETE SAMPLE
M2516-01	1-5-TP	2	1.14	9.75	9.45	96.5	
M2516-02	1-5-TP-EPH-2	3	1.14	9.94	9.51	95.1	
M2517-01	METRO-SP	4	1.16	9.63	9.47	98.1	
M2517-02	METRO-SP-EPH-2	5	1.15	9.75	9.57	97.9	
M2518-01	115	6	1.14	9.66	8.61	87.7	
M2518-02	117	7	1.18	9.55	6.66	65.5	
M2519-01	26878	8	1.15	9.96	8.9	88.0	
M2519-02	26878-EPH-2	9	1.18	9.92	8.58	84.7	
M2520-01	WC-01-2021-05-24	17	1.15	9.96	8.34	81.6	
M2521-01	HR-01-052521	10	1.13	9.91	9.35	93.6	
M2521-02	HR-01-052521-EPH-2	11	1.18	9.66	9.37	96.6	
M2521-03	HR-03-052521	12	1.13	9.87	9.7	98.1	
M2521-04	HR-03-052521-EPH-2	13	1.12	9.64	9.62	99.8	
M2521-04D	HR-03-052521-EPH-2DUP	14	1.12	9.66	9.64	99.8	
M2525-01	TS-5	18	1.12	9.98	7.96	77.2	
M2526-01	OF-3-20210525	19	1.12	9.66	5.81	54.9	
M2526-02	M2526-01MS	20	1.12	9.66	5.81	54.9	
M2526-03	M2526-01MSD	21	1.12	9.66	5.81	54.9	
M2526-04	OF-4-20210525	22	1.13	9.95	4.96	43.4	
M2526-05	DUP03-20210525	23	1.14	9.97	5.1	44.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$