



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114636

Slope : 98.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.2	4.01	05/25/2021	16:26
2	CAL2	1	Water	NA	NA	21.2	7.00	05/25/2021	16:27
3	CAL3	1	Water	NA	NA	21.2	10.02	05/25/2021	16:28
4	ICV	1	Water	NA	NA	21.2	7.00	05/25/2021	16:29
5	CCV1	1	Water	NA	NA	21.4	2.01	05/25/2021	16:30
6	M2514-01	1	Water	NA	NA	20.4	11.86	05/25/2021	16:32
7	M2514-02	1	Water	NA	NA	20.3	11.99	05/25/2021	16:33
8	M2514-03	1	Water	NA	NA	20.3	11.87	05/25/2021	16:34
9	M2514-04	1	Water	NA	NA	20.6	11.98	05/25/2021	16:35
10	M2514-05	1	Water	NA	NA	20.3	11.50	05/25/2021	16:36
11	M2517-03	1	Water	NA	NA	21.8	7.49	05/25/2021	16:37
12	M2518-03	1	Water	NA	NA	21.9	7.50	05/25/2021	16:38
13	M2518-04	1	Water	NA	NA	22.2	7.38	05/25/2021	16:39
14	M2518-06	1	Water	NA	NA	21.3	6.08	05/25/2021	16:40
15	M2518-07	1	Water	NA	NA	21.4	6.14	05/25/2021	16:42
16	CCV2	1	Water	NA	NA	21.1	12.02	05/25/2021	16:43
17	M2518-08	1	Water	NA	NA	21.6	6.19	05/25/2021	16:44
18	M2518-11	1	Water	NA	NA	21.2	6.94	05/25/2021	16:45
19	M2520-03	1	Water	NA	NA	21.2	7.29	05/25/2021	16:46
20	M2520-03DUP	1	Water	NA	NA	21.3	7.30	05/25/2021	16:47

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
21	CCV3	1	Water	NA	NA	21.3	2.01	05/25/2021	16:49

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH M2514 WorkList ID : 149212 Department : Wet-Chemistry Date : 05-25-2021 14:29:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/25/2021	Water	M2514-01	pH	Cool 4 deg C	PSEG03	F61	TOTE-A	05/25/2021	9040C
05/25/2021	Water	M2514-02	pH	Cool 4 deg C	PSEG03	F61	TOTE-B	05/25/2021	9040C
05/25/2021	Water	M2514-03	pH	Cool 4 deg C	PSEG03	F61	TOTE-C	05/25/2021	9040C
05/25/2021	Water	M2514-04	pH	Cool 4 deg C	PSEG03	F61	TOTE-D	05/25/2021	9040C
05/25/2021	Water	M2514-05	pH	Cool 4 deg C	PSEG03	F61	TOTE-E	05/25/2021	9040C
05/25/2021	Water	M2517-03	pH	Cool 4 deg C	PSEG03	F61	1917	05/25/2021	9040C
05/25/2021	Water	M2518-03	pH	Cool 4 deg C	PSEG03	F61	118	05/25/2021	9040C
05/25/2021	Water	M2518-04	pH	Cool 4 deg C	PSEG03	F61	119	05/25/2021	9040C
05/25/2021	Water	M2518-06	pH	Cool 4 deg C	PSEG03	F61	112	05/25/2021	9040C
05/25/2021	Water	M2518-07	pH	Cool 4 deg C	PSEG03	F61	113	05/25/2021	9040C
05/25/2021	Water	M2518-08	pH	Cool 4 deg C	PSEG03	F61	114	05/25/2021	9040C
05/25/2021	Water	M2518-11	pH	Cool 4 deg C	PSEG03		116	05/25/2021	9040C
05/24/2021	Water	M2520-03	pH	Cool 4 deg C	AECO11	D61	WC-03-2021-05-24	05/24/2021	9040C

Date/Time 05/25/21 16:25
 Raw Sample Received by: Jol (04:00pm)
 Raw Sample Relinquished by: FS (Webster)

Date/Time 05/25/21 17:40
 Raw Sample Received by: FS (Webster)
 Raw Sample Relinquished by: Jol (04:00pm)