



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114641

Ambient Barometric Pressure (mmHg): 745.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	81.5	05/26/2021	09:20
2	M2512-02		1	8	80.00	176.0	176.9	05/26/2021	09:45
3	M2512-02DUP		1	8	82.00	179.6	180.5	05/26/2021	10:10
4	M2512-05		1	9	85.00	185.0	185.9	05/26/2021	10:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114641

WorkList Name : fp52621-1 WorkList ID : 149237 Department : Wet-Chemistry Date : 05-26-2021 08:55:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/31/2021	Solid	M2512-02	Flash Point	Cool 4 deg C	PSEG03	F11	COMP-1-3	05/25/2021	1010A
05/31/2021	Solid	M2512-05	Flash Point	Cool 4 deg C	PSEG03	F11	COMP-4-6	05/25/2021	1010A

Date/Time 05/26/21 09:00
 Raw Sample Received by: KS (Webchem)
 Raw Sample Relinquished by: REC Ext Lab

Date/Time 05/26/21 10:45
 Raw Sample Received by: REC Ext Lab
 Raw Sample Relinquished by: KS (Webchem)