



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114642

Ambient Barometric Pressure (mmHg): 745.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	81.5	05/26/2021	09:20
2	M2518-05		1	9	100.00	>212.0	>212.0	05/26/2021	11:05
3	M2518-05DUP		1	9	100.00	>212.0	>212.0	05/26/2021	11:35
4	M2518-09		1	8	100.00	>212.0	>212.0	05/26/2021	12:10
5	M2518-11		1	9	100.00	>212.0	>212.0	05/26/2021	12:40
6	M2520-03		1	9	100.00	>212.0	>212.0	05/26/2021	13:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114642

WorkList Name : fp52621 WorkList ID : 149234 Department : Wet-Chemistry Date : 05-26-2021 08:54:17

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/01/2021	Water	M2518-05	Flash Point	Cool 4 deg C	PSEG03	F61	118-119	05/25/2021	1010A
06/01/2021	Water	M2518-09	Flash Point	Cool 4 deg C	PSEG03	F61	112-114	05/25/2021	1010A
06/01/2021	Water	M2518-11	Flash Point	Cool 4 deg C	PSEG03	116		05/25/2021	1010A
05/28/2021	Water	M2520-03	Flash Point	Cool 4 deg C	AECO11	D61	WC-03-2021-05-24	05/24/2021	1010A

Date/Time 05/26/21 09:00
 Raw Sample Received by: KS (Wetche)
 Raw Sample Relinquished by: RC (C-x+G-b)

Date/Time 05/26/21 13:15
 Raw Sample Received by: RC (C-x+G-b)
 Raw Sample Relinquished by: KS (Wetche)