


LB114649

SFAM01.1

INSTRUMENT ID: CY4



Linear

A= 0.0000e+000

B= 2.4262e-004 slope

C= -4.0140e-002 y-intercept

Rho= 0.9999322

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.001	0.001	171	0.000	171					
0.05	0.050					204					
0.20	0.200	0.189	-0.011	943	0.0 %	943					
2.50	2.500	2.511	0.011	10515	0.0 %	10515					
5.00	5.000	5.048	0.048	20970	0.0 %	20970					
7.50	7.500	7.415	-0.085	30727	0.0 %	30727					
10.0	10.000	10.036	0.036	41531	0.0 %	41531					

90D
=60110

LB114649
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	171	-		0 SFAM01.1	PPB	5/26/2021 14:04	Std
	0.2 S01	943	-		0.2 SFAM01.1	PPB	5/26/2021 14:06	Std
	2.5 S02	10515	-		2.5 SFAM01.1	PPB	5/26/2021 14:09	Std
	5 S03	20970	-		5 SFAM01.1	PPB	5/26/2021 14:11	Std
	7.5 S04	30727	-		7.5 SFAM01.1	PPB	5/26/2021 14:13	Std
	10 S05	41531	-		10 SFAM01.1	PPB	5/26/2021 14:21	Std
ICV034	ICV034	15470	3.7133 -		SFAM01.1	PPB	5/26/2021 14:23	SMPL
ICB034	ICB034	76	-0.0217 -		SFAM01.1	PPB	5/26/2021 14:26	SMPL
CCV070	CCV070	20941	5.0407 -		SFAM01.1	PPB	5/26/2021 14:28	SMPL
CCB070	CCB070	76	-0.0217 -		SFAM01.1	PPB	5/26/2021 14:30	SMPL
PB136664BL	PBS664	77	-0.0215 -		SFAM01.1	PPB	5/26/2021 14:35	SMPL
M2400-19	X3375	20902	5.0312 -		SFAM01.1	PPB	5/26/2021 14:37	SMPL
M2400-20	X3375D	20592	4.956 -		SFAM01.1	PPB	5/26/2021 14:42	SMPL
PB136666BL	PBS666	173	0.0018 -		SFAM01.1	PPB	5/26/2021 14:47	SMPL
M2412-01	MHOAY4	525	0.0872 -		SFAM01.1	PPB	5/26/2021 14:49	SMPL
M2412-02	MHOAY5	1308	0.2772 -		SFAM01.1	PPB	5/26/2021 14:51	SMPL
M2412-03	MHOAY6	756	0.1433 -		SFAM01.1	PPB	5/26/2021 14:53	SMPL
M2412-04	MHOAY7	1065	0.2183 -		SFAM01.1	PPB	5/26/2021 14:56	SMPL
M2412-05	MHOAY7D	1171	0.244 -		SFAM01.1	PPB	5/26/2021 14:58	SMPL
M2412-06	MHOAY7S	11472	2.7432 -		SFAM01.1	PPB	5/26/2021 15:00	SMPL
PB136667BL	PBS667	-27	-0.0467 -		SFAM01.1	PPB	5/26/2021 15:02	SMPL
M2477-01	MA0BE9	240	0.0181 -		SFAM01.1	PPB	5/26/2021 15:05	SMPL
M2477-02	MA0BE9D	185	0.0047 -		SFAM01.1	PPB	5/26/2021 15:07	SMPL
M2477-03	MA0BE9S	11247	2.6887 -		SFAM01.1	PPB	5/26/2021 15:09	SMPL
M2477-04	MA0BF0	1358	0.2893 -		SFAM01.1	PPB	5/26/2021 15:11	SMPL
M2477-05	MA0BF1	209	0.0106 -		SFAM01.1	PPB	5/26/2021 15:14	SMPL
CCV071	CCV071	21035	5.0635 -		SFAM01.1	PPB	5/26/2021 15:16	SMPL
CCB071	CCB071	67	-0.0239 -		SFAM01.1	PPB	5/26/2021 15:18	SMPL
M2477-06	MA0BF2	292	0.0307 -		SFAM01.1	PPB	5/26/2021 15:21	SMPL
M2477-07	MA0BF3	282	0.0283 -		SFAM01.1	PPB	5/26/2021 15:23	SMPL
M2477-08	MA0BF4	239	0.0178 -		SFAM01.1	PPB	5/26/2021 15:25	SMPL
M2477-09	MA0BF5	111	-0.0132 -		SFAM01.1	PPB	5/26/2021 15:27	SMPL
M2477-10	MA0BE8	421	0.062 -		SFAM01.1	PPB	5/26/2021 15:30	SMPL
M2477-11	MA0BF6	166	0.0001 -		SFAM01.1	PPB	5/26/2021 15:32	SMPL
M2477-12	MA0BF7	177	0.0028 -		SFAM01.1	PPB	5/26/2021 15:34	SMPL
M2477-13	MA0BF8	772	0.1472 -		SFAM01.1	PPB	5/26/2021 15:36	SMPL
M2477-14	MA0BF9	618	0.1098 -		SFAM01.1	PPB	5/26/2021 15:39	SMPL
M2477-15	MA0BG0	1407	0.3012 -		SFAM01.1	PPB	5/26/2021 15:41	SMPL
M2477-16	MA0BG3	18159	4.3657 -		SFAM01.1	PPB	5/26/2021 15:45	SMPL
M2477-17	MA0BG4	446	0.0681 -		SFAM01.1	PPB	5/26/2021 15:50	SMPL
M2400-21	X3375S	30428	7.3424 -		SFAM01.1	PPB	5/26/2021 15:52	SMPL
CCV072	CCV072	20009	4.8145 -		SFAM01.1	PPB	5/26/2021 15:55	SMPL
CCB072	CCB072	41	-0.0302 -		SFAM01.1	PPB	5/26/2021 15:57	SMPL