



Analytical Summary Report

Analysis Method: 9040C

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114655

Slope : 98.8

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	05/26/2021	16:20
2	CAL2	1	Water	NA	NA	21.4	7.00	05/26/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.02	05/26/2021	16:22
4	ICV	1	Water	NA	NA	21.3	7.00	05/26/2021	16:23
5	CCV1	1	Water	NA	NA	21.4	2.01	05/26/2021	16:24
6	M2529-01	1	Water	NA	NA	21.5	6.17	05/26/2021	16:25
7	M2529-01DUP	1	Water	NA	NA	21.5	6.15	05/26/2021	16:26
8	M2529-02	1	Water	NA	NA	21.5	6.89	05/26/2021	16:27
9	M2529-03	1	Water	NA	NA	21.9	6.74	05/26/2021	16:28
10	M2529-04	1	Water	NA	NA	22.5	9.14	05/26/2021	16:29
11	M2532-01	1	Water	NA	NA	24.0	7.76	05/26/2021	16:30
12	CCV2	1	Water	NA	NA	21.2	12.02	05/26/2021	16:32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph52621 WorkList ID : 149259 Department : Wet-Chemistry Date : 05-26-2021 13:05:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/26/2021	Water	M2529-01	pH	Cool 4 deg C	PSEG03	E61	31046	05/26/2021	9040C
05/26/2021	Water	M2529-02	pH	Cool 4 deg C	PSEG03	E61	41791	05/26/2021	9040C
05/26/2021	Water	M2529-03	pH	Cool 4 deg C	PSEG03	E61	41822	05/26/2021	9040C
05/26/2021	Water	M2529-04	pH	Cool 4 deg C	PSEG03	E61	41815	05/26/2021	9040C
05/26/2021	Water	M2532-01	pH	Cool 4 deg C	TULL02	D61	MH-N-(MH-10.29)	05/26/2021	9040C

Date/Time 05/26/21 16:10
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: SD (Wetchem)

Date/Time 05/26/21 16:35
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: KS (Wetchem)