



Analytical Summary Report

Analysis Method: 9045D

Analyst By : ketankumar

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114656

Slope : 98.8

BalanceID:

pH Meter ID : WC pH Meter-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.3	4.01	05/26/2021	16:20
2	CAL2	1	Water	NA	NA	21.4	7.00	05/26/2021	16:21
3	CAL3	1	Water	NA	NA	21.2	10.02	05/26/2021	16:22
4	ICV	1	Water	NA	NA	21.3	7.00	05/26/2021	16:23
5	CCV1	1	Water	NA	NA	21.3	2.01	05/26/2021	16:35
6	M2539-01	1	Solid	20	20	25.8	7.16	05/26/2021	16:36
7	M2539-01DUP	1	Solid	20	20	25.8	7.15	05/26/2021	16:37
8	M2542-01	1	Solid	20	20	25.8	6.99	05/26/2021	16:38
9	CCV2	1	Water	NA	NA	21.6	12.01	05/26/2021	16:39

WORKLIST(Hardcopy Internal Chain)

LB 114656

WorkList Name : ph52621-1 WorkList ID : 149263 Department : Wet-Chemistry Date : 05-26-2021 16:14:36

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/26/2021	Solid	M2539-01	pH	Cool 4 deg C	TULL02	D62	FK-PS-01-073	05/26/2021	9045D
05/26/2021	Solid	M2542-01	pH	Cool 4 deg C	TULL02	D61	FK-PS-01-074	05/26/2021	9045D

Date/Time 05/26/21 16:20
 Raw Sample Received by: KS (Welchen)
 Raw Sample Relinquished by: JPC (Welchen)

Date/Time 05/26/21 16:40
 Raw Sample Received by: JPC (Welchen)
 Raw Sample Relinquished by: KS (Welchen)