


LB114660

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.059	0.059	39	0.000	39					
0.2	0.200	0.224	0.024	647	0.0 %	647					
2.5	2.500	2.410	-0.090	8720	0.0 %	8720					
5.0	5.000	4.977	-0.023	18202	0.0 %	18202					
7.5	7.500	7.494	-0.006	27500	0.0 %	27500					
10.0	10.000	10.037	0.037	36889	0.0 %	36889					

Handwritten notes on the right side of the table:

12

-4

0000

LB114660 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 May 2021 11:00:40

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	39	-	0.0000	7470A	PPB	27 May 2021 11:03:15	S	Std
0.2 - 1	50.2	647	-	0.2000	7470A	PPB	27 May 2021 11:05:32	S	Std
2.5 - 1	52.5	8720	-	2.5000	7470A	PPB	27 May 2021 11:07:49	S	Std
5.0 - 1	55	18202	-	5.0000	7470A	PPB	27 May 2021 11:10:06	S	Std
7.5 - 1	57.5	27500	-	7.5000	7470A	PPB	27 May 2021 11:12:22	S	Std
10.0 - 1	31.8	36889	-	10.0000	7470A	PPB	27 May 2021 11:14:37	S	Std
ICV40 - 1	ICV40	13154	3.6101	-	7470A	PPB	27 May 2021 11:19:59	U	SMPL
ICB40 - 1	ICB40	-120	0.0160	-	7470A	PPB	27 May 2021 11:22:14	U	SMPL
CCV73 - 1	CCV73	17566	4.8047	-	7470A	PPB	27 May 2021 11:32:14	U	SMPL
CCB73 - 1	CCB73	-151	0.0076	-	7470A	PPB	27 May 2021 11:34:28	U	SMPL
CRA - 1	CRA	552	0.1980	-	7470A	PPB	27 May 2021 11:36:44	U	SMPL
HighStd - 1	HighStd	39393	10.7145	-	7470A	PPB	27 May 2021 11:40:15	U	SMPL
ChkStd - 1	ChkStd	27226	7.4202	-	7470A	PPB	27 May 2021 11:42:29	U	SMPL
PB136660BL - 1	PBW	-159	0.0054	-	7470A	PPB	27 May 2021 11:44:46	U	SMPL
PB136660BS - 1	LCSW	13540	3.7146	-	7470A	PPB	27 May 2021 11:47:02	U	SMPL
M2512-02 - 1	COMP-1-3	2605	0.7538	-	7470A	PPB	27 May 2021 11:49:18	U	SMPL
M2512-05 - 1	COMP-4-6	2432	0.7070	-	7470A	PPB	27 May 2021 11:51:35	U	SMPL
M2518-01 - 1	115	63	0.0656	-	7470A	PPB	27 May 2021 11:53:52	U	SMPL
M2518-02 - 1	117	57	0.0639	-	7470A	PPB	27 May 2021 11:56:09	U	SMPL
M2520-01 - 1	WC-01-2021-05-24	76	0.0691	-	7470A	PPB	27 May 2021 11:58:24	U	SMPL
CCV74 - 1	CCV74	18356	5.0185	-	7470A	PPB	27 May 2021 12:00:39	U	SMPL
CCB74 - 1	CCB74	-120	0.0160	-	7470A	PPB	27 May 2021 12:02:54	U	SMPL
M2520-01DUP - 1	WC-01-2021-05-24DUP	67	0.0666	-	7470A	PPB	27 May 2021 12:05:11	U	SMPL
M2520-01MS - 1	WC-01-2021-05-24MS	12859	3.5302	-	7470A	PPB	27 May 2021 12:07:26	U	SMPL
M2520-01MSD - 1	WC-01-2021-05-24MSD	13175	3.6157	-	7470A	PPB	27 May 2021 12:09:41	U	SMPL
PB136661BL - 1	PBW	-60	0.0322	-	7470A	PPB	27 May 2021 12:11:57	U	SMPL
PB136661BS - 1	LCSW	12531	3.4414	-	7470A	PPB	27 May 2021 12:14:13	U	SMPL
M2517-03 - 1	1917	84	0.0712	-	7470A	PPB	27 May 2021 12:16:29	U	SMPL
M2518-05 - 1	118-119	80	0.0702	-	7470A	PPB	27 May 2021 12:18:45	U	SMPL
M2520-03 - 1	WC-03-2021-05-24	63	0.0656	-	7470A	PPB	27 May 2021 12:21:02	U	SMPL
M2520-03DUP - 1	WC-03-2021-05-24DUP	93	0.0737	-	7470A	PPB	27 May 2021 12:23:19	U	SMPL
M2520-03MS - 1	WC-03-2021-05-24MS	15486	4.2415	-	7470A	PPB	27 May 2021 12:25:35	U	SMPL
CCV75 - 1	CCV75	18721	5.1174	-	7470A	PPB	27 May 2021 12:27:50	U	SMPL
CCB75 - 1	CCB75	-67	0.0304	-	7470A	PPB	27 May 2021 12:30:05	U	SMPL
M2520-03MSD - 1	WC-03-2021-05-24MSD	14860	4.0720	-	7470A	PPB	27 May 2021 12:32:21	U	SMPL
PB136662BL - 1	PBW	-27	0.0412	-	7470A	PPB	27 May 2021 12:34:36	U	SMPL
PB136662BS - 1	LCSW	12166	3.3426	-	7470A	PPB	27 May 2021 12:36:53	U	SMPL
M2518-09 - 1	112-114	22	0.0545	-	7470A	PPB	27 May 2021 12:39:08	U	SMPL
M2518-09DUP - 1	112-114DUP	122	0.0815	-	7470A	PPB	27 May 2021 12:41:23	U	SMPL
M2518-09MS - 1	112-114MS	14861	4.0722	-	7470A	PPB	27 May 2021 12:43:39	U	SMPL
M2518-09MSD - 1	112-114MSD	14968	4.1012	-	7470A	PPB	27 May 2021 12:45:55	U	SMPL
M2518-11 - 1	116	2786	0.8028	-	7470A	PPB	27 May 2021 12:48:12	U	SMPL
M2526-06 - 1	FB02-20210525	-18	0.0436	-	7470A	PPB	27 May 2021 12:50:29	U	SMPL
M2529-04 - 1	41815	704	0.2391	-	7470A	PPB	27 May 2021 12:52:44	U	SMPL
CCV76 - 1	CCV76	17289	4.7296	-	7470A	PPB	27 May 2021 12:54:59	U	SMPL
CCB76 - 1	CCB76	-117	0.0168	-	7470A	PPB	27 May 2021 12:57:16	U	SMPL
PB136660TB - 1	PB136645TB	65	0.0661	-	7470A	PPB	27 May 2021 12:59:32	U	SMPL
PB136661TB - 1	PB136644TB	72	0.0680	-	7470A	PPB	27 May 2021 13:01:47	U	SMPL
M2520-01LX5 - 1		16	0.0528	-	7470A	PPB	27 May 2021 13:04:02	U	SMPL
M2520-01A - 1		14573	3.9943	-	7470A	PPB	27 May 2021 13:06:18	U	SMPL
M2520-03LX5 - 1		-66	0.0306	-	7470A	PPB	27 May 2021 13:08:33	U	SMPL
M2520-03A - 1		14641	4.0127	-	7470A	PPB	27 May 2021 13:10:49	U	SMPL
M2518-09LX5 - 1		-52	0.0344	-	7470A	PPB	27 May 2021 13:13:05	U	SMPL
M2518-09A - 1		14567	3.9926	-	7470A	PPB	27 May 2021 13:15:22	U	SMPL
CCV77 - 1	CCV77	18138	4.9595	-	7470A	PPB	27 May 2021 13:17:39	U	SMPL
CCB77 - 1	CCB77	-98	0.0220	-	7470A	PPB	27 May 2021 13:19:55	U	SMPL