



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114662

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.6	05/27/2021	10:20
2	M2529-01		1	9	100.00	>212.0	>212.0	05/27/2021	10:45
3	M2529-01DUP		1	9	100.00	>212.0	>212.0	05/27/2021	11:10
4	M2529-02		1	8	100.00	>212.0	>212.0	05/27/2021	11:40
5	M2529-03		1	9	100.00	>212.0	>212.0	05/27/2021	12:10
6	M2529-04		1	8	100.00	>212.0	>212.0	05/27/2021	12:35
7	M2532-01		1	9	100.00	>212.0	>212.0	05/27/2021	12:58

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP52721

WorkList ID : 149289

Department : Wet-Chemistry

Date : 05-27-2021 09:29:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
/02/2021	Water	M2529-01	Flash Point	Cool 4 deg C	PSEG03	E61	31046	05/26/2021	1010A
/02/2021	Water	M2529-02	Flash Point	Cool 4 deg C	PSEG03	E61	41791	05/26/2021	1010A
/02/2021	Water	M2529-03	Flash Point	Cool 4 deg C	PSEG03	E61	41822	05/26/2021	1010A
/02/2021	Water	M2529-04	Flash Point	Cool 4 deg C	PSEG03	E61	41815	05/26/2021	1010A
/02/2021	Water	M2532-01	Flash Point	Cool 4 deg C	TULL02	D61	MH-N-(MH-10.29)	05/26/2021	1010A

LB114662