

LB114666


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

47370



A= 0.0000e+000

B= 2.0954e-004

C= -5.9697e-003

Rho= 0.9998387

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	405
0.0	0.000	0.012	0.012	86	0.000	86					1
0.2	0.200	0.204	0.004	1000	0.0 %	1000					2
2.5	2.500	2.450	-0.050	11721	0.0 %	11721					-2
5.0	5.000	4.984	-0.016	23813	0.0 %	23813					0
7.5	7.500	7.631	0.131	36444	0.0 %	36444					2
10.0	10.000	9.920	-0.080	47370	0.0 %	47370					1

LB114666

Method: 7471B

Operator: Admin

Date of Analysis: 27 May 2021 13:48:57

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	86	-	0.0000	7471B	PPB	27 May 2021 13:50:48	S	Std
0.2 - 1	510	1000	-	0.2000	7471B	PPB	27 May 2021 13:53:05	S	Std
2.5 - 1	52.5	11721	-	2.5000	7471B	PPB	27 May 2021 13:55:22	S	Std
5.0 - 1	55	23813	-	5.0000	7471B	PPB	27 May 2021 13:57:39	S	Std
7.5 - 1	57.5	36444	-	7.5000	7471B	PPB	27 May 2021 13:59:55	S	Std
10.0 - 1	510	47370	-	10.0000	7471B	PPB	27 May 2021 14:02:12	S	Std
ICV41 - 1	ICV41	18577	3.8867	-7471B	PPB	27 May 2021 14:04:50	U	SMPL	
ICB41 - 1	ICB41	-83	-0.0234	-7471B	PPB	27 May 2021 14:07:05	U	SMPL	
CCV78 - 1	CCV78	24106	5.0452	-7471B	PPB	27 May 2021 14:09:20	U	SMPL	
CCB78 - 1	CCB78	-93	-0.0255	-7471B	PPB	27 May 2021 14:11:35	U	SMPL	
CRA - 1	CRA	944	0.1918	-7471B	PPB	27 May 2021 14:13:51	U	SMPL	
HighStd - 1	HighStd	44537	9.3263	-7471B	PPB	27 May 2021 14:16:06	U	SMPL	
ChkStd - 1	ChkStd	32197	6.7406	-7471B	PPB	27 May 2021 14:18:21	U	SMPL	
PB136678BL - 1	PBS	-149	-0.0372	-7471B	PPB	27 May 2021 14:20:39	U	SMPL	
PB136678BS - 1	LCSS	17663	3.6951	-7471B	PPB	27 May 2021 14:28:20	U	SMPL	
M2512-01 - 1	COMP-1-3	800	0.1617	-7471B	PPB	27 May 2021 14:30:36	U	SMPL	
M2512-04 - 1	COMP-4-6	908	0.1843	-7471B	PPB	27 May 2021 14:32:54	U	SMPL	
M2516-01 - 1	1-5-TP	241	0.0445	-7471B	PPB	27 May 2021 14:35:10	U	SMPL	
M2517-01 - 1	METRO-SP	1553	0.3194	-7471B	PPB	27 May 2021 14:37:27	U	SMPL	
M2519-01 - 1	26878	10265	2.1450	-7471B	PPB	27 May 2021 14:39:42	U	SMPL	
CCV79 - 1	CCV79	23401	4.8975	-7471B	PPB	27 May 2021 14:41:57	U	SMPL	
CCB79 - 1	CCB79	-45	-0.0154	-7471B	PPB	27 May 2021 14:44:11	U	SMPL	
M2521-01 - 1	HR-01-052521	2452	0.5078	-7471B	PPB	27 May 2021 14:46:27	U	SMPL	
M2521-03 - 1	HR-03-052521	9065	1.8935	-7471B	PPB	27 May 2021 14:48:42	U	SMPL	
M2525-01 - 1	TS-5	2446	0.5066	-7471B	PPB	27 May 2021 14:50:58	U	SMPL	
M2526-01 - 1	OF-3-20210525	16674	3.4879	-7471B	PPB	27 May 2021 14:53:14	U	SMPL	
M2526-01DUP - 1	OF-3-20210525DUP	15810	3.3069	-7471B	PPB	27 May 2021 14:55:29	U	SMPL	
M2526-02 - 1	M2536-01MS	33236	6.9583	-7471B	PPB	27 May 2021 14:57:54	U	SMPL	
M2526-03 - 1	M2536-01MSD	34100	7.1394	-7471B	PPB	27 May 2021 15:00:10	U	SMPL	
M2526-04 - 1	OF-4-20210525	5986	1.2483	-7471B	PPB	27 May 2021 15:02:28	U	SMPL	
M2526-05 - 1	DUP03-20210525	8463	1.7674	-7471B	PPB	27 May 2021 15:04:46	U	SMPL	
M2531-01 - 1	SU-04-052621	1496	0.3075	-7471B	PPB	27 May 2021 15:07:01	U	SMPL	
CCV80 - 1	CCV80	23106	4.8357	-7471B	PPB	27 May 2021 15:09:17	U	SMPL	
CCB80 - 1	CCB80	-17	-0.0095	-7471B	PPB	27 May 2021 15:11:32	U	SMPL	
M2536-01 - 1	FK-BPM-01-C7-C7A-D7-D7A	947	0.1925	-7471B	PPB	27 May 2021 15:13:48	U	SMPL	
M2537-01 - 1	FK-BPM-04-016-ABCD	603	0.1204	-7471B	PPB	27 May 2021 15:16:03	U	SMPL	
M2538-01 - 1	FK-BPM-03-022-ABCD	1830	0.3775	-7471B	PPB	27 May 2021 15:18:18	U	SMPL	
M2539-01 - 1	FK-PS-01-073	1747	0.3601	-7471B	PPB	27 May 2021 15:20:33	U	SMPL	
M2542-01 - 1	FK-PS-01-074	2066	0.4269	-7471B	PPB	27 May 2021 15:22:49	U	SMPL	
M2526-01LX5 - 1		3147	0.6535	-7471B	PPB	27 May 2021 15:25:06	U	SMPL	
M2526-01A - 1		34573	7.2385	-7471B	PPB	27 May 2021 15:27:22	U	SMPL	
CCV81 - 1	CCV81	23074	4.8290	-7471B	PPB	27 May 2021 15:32:07	U	SMPL	
CCB81 - 1	CCB81	-25	-0.0112	-7471B	PPB	27 May 2021 15:34:23	U	SMPL	