

LB114 674


SFAM01.1

INSTRUMENT ID: LV2

Linear

 μ Abs.:

36647



A= 0.0000e+000

B= 2.7188e-004 slope

C= -3.7603e-002 y-intercept

Rho= 0.9998987

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Valid
0.00	0.000	-0.004	-0.004	122	0.000	122					=
0.05	0.050					204					
0.20	0.200	0.198	-0.002	865	0.0 %	865					-1
2.50	2.500	2.473	-0.027	9235	0.0 %	9235					-1
5.00	5.000	5.007	0.007	18553	0.0 %	18553					0
7.50	7.500	7.601	0.101	28097	0.0 %	28097					1
10.0	10.000	9.926	-0.074	36647	0.0 %	36647					-1

LB114674
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	122	-		0 SFAM01.1	PPB	5/27/2021 15:55	Std
	0.2 S01	865	-		0.2 SFAM01.1	PPB	5/27/2021 15:58	Std
	2.5 S02	9235	-		2.5 SFAM01.1	PPB	5/27/2021 16:00	Std
	5 S03	18553	-		5 SFAM01.1	PPB	5/27/2021 16:02	Std
	7.5 S04	28097	-		7.5 SFAM01.1	PPB	5/27/2021 16:05	Std
	10 S05	36647	-		10 SFAM01.1	PPB	5/27/2021 16:07	Std
ICV035	ICV035	14420	3.8829	-	SFAM01.1	PPB	5/27/2021 16:10	SMPL
ICB035	ICB035	40	-0.0267	-	SFAM01.1	PPB	5/27/2021 16:12	SMPL
CCV082	CCV082	18034	4.8654	-	SFAM01.1	PPB	5/27/2021 16:14	SMPL
CCB082	CCB082	-2	-0.0381	-	SFAM01.1	PPB	5/27/2021 16:16	SMPL
PB136692BL	PBW692	12	-0.0343	-	SFAM01.1	PPB	5/27/2021 16:19	SMPL
M2393-01	MAOAM4	28	-0.03	-	SFAM01.1	PPB	5/27/2021 16:21	SMPL
M2393-02	MAOAP8	32	-0.0289	-	SFAM01.1	PPB	5/27/2021 16:23	SMPL
M2393-03	MAO AQ6	36	-0.0278	-	SFAM01.1	PPB	5/27/2021 16:25	SMPL
M2393-05	MAOAY8	230	0.0249	-	SFAM01.1	PPB	5/27/2021 16:30	SMPL
M2393-06	MAOAZ0	33	-0.0286	-	SFAM01.1	PPB	5/27/2021 16:32	SMPL
M2393-07	MAOAZ2	78	-0.0164	-	SFAM01.1	PPB	5/27/2021 16:35	SMPL
M2393-08	MAOAM6	34	-0.0284	-	SFAM01.1	PPB	5/27/2021 16:37	SMPL
M2393-09	MAOAN4	43	-0.0259	-	SFAM01.1	PPB	5/27/2021 16:39	SMPL
M2393-10	MAO AQ8	67	-0.0194	-	SFAM01.1	PPB	5/27/2021 16:41	SMPL
M2393-11	MAOAR2	44	-0.0256	-	SFAM01.1	PPB	5/27/2021 16:44	SMPL
M2393-12	MAO AQ4	44	-0.0256	-	SFAM01.1	PPB	5/27/2021 16:46	SMPL
M2393-13	MAOAW8	96	-0.0115	-	SFAM01.1	PPB	5/27/2021 16:48	SMPL
M2393-14	MAOAX6	32	-0.0289	-	SFAM01.1	PPB	5/27/2021 16:50	SMPL
M2393-15	MAOAM8	33	-0.0286	-	SFAM01.1	PPB	5/27/2021 16:53	SMPL
M2393-16	MAOAS0	45	-0.0254	-	SFAM01.1	PPB	5/27/2021 16:55	SMPL
M2393-17	MAOAS0D	36	-0.0278	-	SFAM01.1	PPB	5/27/2021 16:57	SMPL
M2393-18	MAOAS0S	3876	1.0162	-	SFAM01.1	PPB	5/27/2021 16:59	SMPL
M2393-19	MAOAS4	12	-0.0343	-	SFAM01.1	PPB	5/27/2021 17:02	SMPL
CCV083	CCV083	18509	4.9946	-	SFAM01.1	PPB	5/27/2021 17:04	SMPL
CCB083	CCB083	23	-0.0314	-	SFAM01.1	PPB	5/27/2021 17:06	SMPL
M2393-20	MAOAS6	20	-0.0322	-	SFAM01.1	PPB	5/27/2021 17:09	SMPL
M2393-21	MAOAX0	35	-0.0281	-	SFAM01.1	PPB	5/27/2021 17:11	SMPL
PB136693BL	PBW693	18	-0.0327	-	SFAM01.1	PPB	5/27/2021 17:13	SMPL
M2394-01	MAOAM5	14	-0.0338	-	SFAM01.1	PPB	5/27/2021 17:15	SMPL
M2394-02	MAOAP9	-9	-0.0401	-	SFAM01.1	PPB	5/27/2021 17:18	SMPL
M2394-03	MAO AQ7	46	-0.0251	-	SFAM01.1	PPB	5/27/2021 17:20	SMPL
M2394-04	MAOAY1	15539	4.1871	-	SFAM01.1	PPB	5/27/2021 17:25	SMPL
M2394-05	MAOAY9	-34	-0.0468	-	SFAM01.1	PPB	5/27/2021 17:27	SMPL
M2394-06	MAOAZ1	21	-0.0319	-	SFAM01.1	PPB	5/27/2021 17:29	SMPL
M2394-07	MAOAZ3	35	-0.0281	-	SFAM01.1	PPB	5/27/2021 17:32	SMPL
M2394-08	MAOAM7	21	-0.0319	-	SFAM01.1	PPB	5/27/2021 17:34	SMPL
M2394-09	MAOAN5	36	-0.0278	-	SFAM01.1	PPB	5/27/2021 17:36	SMPL
M2394-10	MAO AQ9	32	-0.0289	-	SFAM01.1	PPB	5/27/2021 17:38	SMPL
M2394-11	MAOAR3	32	-0.0289	-	SFAM01.1	PPB	5/27/2021 17:41	SMPL
M2394-12	MAO AQ5	33	-0.0286	-	SFAM01.1	PPB	5/27/2021 17:43	SMPL

LB114674

INSTRUMENT ID : CV1

M2394-13	MA0AW9	36	-0.0278 -	SFAM01.1	PPB	5/27/2021 17:45 SMPL
M2394-14	MA0AX7	39	-0.027 -	SFAM01.1	PPB	5/27/2021 17:47 SMPL
M2394-15	MA0AM9	28	-0.03 -	SFAM01.1	PPB	5/27/2021 17:50 SMPL
M2394-16	MA0AS1	41	-0.0265 -	SFAM01.1	PPB	5/27/2021 17:52 SMPL
M2394-17	MA0AS1D	42	-0.0262 -	SFAM01.1	PPB	5/27/2021 17:54 SMPL
CCV084	CCV084	17830	4.81 -	SFAM01.1	PPB	5/27/2021 17:56 SMPL
CCB084	CCB084	-2	-0.0381 -	SFAM01.1	PPB	5/27/2021 17:59 SMPL
M2394-18	MA0AS1S	3804	0.9966 -	SFAM01.1	PPB	5/27/2021 18:01 SMPL
M2394-19	MA0AS5	131	-0.002 -	SFAM01.1	PPB	5/27/2021 18:03 SMPL
M2394-20	MA0AS7	104	-0.0093 -	SFAM01.1	PPB	5/27/2021 18:05 SMPL
M2394-21	MA0AX1	89	-0.0134 -	SFAM01.1	PPB	5/27/2021 18:08 SMPL
M2393-04	MA0AY0	16284	4.3896 -	SFAM01.1	PPB	5/27/2021 18:26 SMPL
CCV085	CCV085	17838	4.8121 -	SFAM01.1	PPB	5/27/2021 18:33 SMPL
CCB085	CCB085	-9	-0.0401 -	SFAM01.1	PPB	5/27/2021 18:35 SMPL