



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB114678

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.6	4.01	05/27/2021	17:36
2	CAL2	1	Water	NA	NA	20.6	7.01	05/27/2021	17:37
3	CAL3	1	Water	NA	NA	20.6	10.02	05/27/2021	17:38
4	ICV	1	Water	NA	NA	20.7	6.99	05/27/2021	17:40
5	CCV1	1	Water	NA	NA	20.6	2.01	05/27/2021	17:42
6	M2551-01	1	Solid	20.02	20	22.2	7.61	05/27/2021	17:45
7	M2551-01DUP	1	Solid	20.03	20	22.1	7.62	05/27/2021	17:46
8	M2551-05	1	Solid	20.04	20	22.1	7.81	05/27/2021	17:48
9	M2551-09	1	Solid	20.02	20	22.3	7.77	05/27/2021	17:49
10	CCV2	1	Water	NA	NA	20.6	12.02	05/27/2021	17:50

203 114678

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph m2551 WorkList ID : 149325 Department : Wet-Chemistry Date : 05-27-2021 17:10:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/27/2021	Solid	M2551-01	pH	Cool 4 deg C	ENTA04	D62	JCSY-SP1	05/27/2021	9045D
05/27/2021	Solid	M2551-05	pH	Cool 4 deg C	ENTA04	D62	JCSY-SP2	05/27/2021	9045D
05/27/2021	Solid	M2551-09	pH	Cool 4 deg C	ENTA04	D62	JCSY-SP3	05/27/2021	9045D

Date/Time 05/27/21 17:20
 Raw Sample Received by: JP (wet Chem)
 Raw Sample Relinquished by: JP (54)

Date/Time 05/27/21 18:25
 Raw Sample Received by: JP (54)
 Raw Sample Relinquished by: JP (wet Chem)