



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114685

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	27.00	80.6	80.3	05/28/2021	09:45
2	M2556-01		1	9	100.00	>212.0	>212.0	05/28/2021	10:10
3	M2556-01DUP		1	9	100.00	>212.0	>212.0	05/28/2021	10:40
4	M2556-02		1	8	100.00	>212.0	>212.0	05/28/2021	11:05

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 114685

WorkList Name : fp52821 WorkList ID : 149317 Department : Wet-Chemistry Date : 05-28-2021 09:08:16

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/03/2021	Water	M2556-01	Flash Point	Cool 4 deg C	DVIR01	D62	MW-8B	05/27/2021	1010A
06/03/2021	Water	M2556-02	Flash Point	Cool 4 deg C	DVIR01	D62	MW-5B	05/27/2021	1010A

Date/Time 05/28/21 09:15
 Raw Sample Received by: KS (Wetchem)
 Raw Sample Relinquished by: RC (Ext Lab)

Date/Time 05/28/21 11:10
 Raw Sample Received by: RC (Ext Lab)
 Raw Sample Relinquished by: KS (Wetchem)