


LB114682

SFAM01.1 INSTRUMENT ID: CVZ

Linear

μ Abs.:

39463



A= 0.0000e+000

B= 2.5376e-004 slope

C= -8.2924e-004 y-intercept

Rho= 0.9999446

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.004	-0.004	-14	0.000	-14					5
0.05	0.050					204					16
0.20	0.200	0.232	0.032	918	0.0 %	918					0
2.50	2.500	2.509	0.009	9891	0.0 %	9891					0
5.00	5.000	4.918	-0.082	19384	0.0 %	19384					-2
7.50	7.500	7.532	0.032	29685	0.0 %	29685					0
10.0	10.000	10.013	0.013	39463	0.0 %	39463					0

LB114688
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	-14	-		0 SFAM01.1	PPB	5/28/2021 12:31	Std
	0.2 S01	918	-		0.2 SFAM01.1	PPB	5/28/2021 12:33	Std
	2.5 S02	9891	-		2.5 SFAM01.1	PPB	5/28/2021 12:35	Std
	5 S03	19384	-		5 SFAM01.1	PPB	5/28/2021 12:37	Std
	7.5 S04	29685	-		7.5 SFAM01.1	PPB	5/28/2021 12:42	Std
	10 S05	39463	-		10 SFAM01.1	PPB	5/28/2021 12:44	Std
ICV036	ICV036	15285	3.8779 -		SFAM01.1	PPB	5/28/2021 12:47	SMPL
ICB036	ICB036	-43	-0.0117 -		SFAM01.1	PPB	5/28/2021 12:49	SMPL
CCV086	CCV086	19233	4.8797 -		SFAM01.1	PPB	5/28/2021 12:52	SMPL
CCB086	CCB086	-69	-0.0183 -		SFAM01.1	PPB	5/28/2021 12:54	SMPL
PB136711BL	PBW711	-8	-0.0029 -		SFAM01.1	PPB	5/28/2021 12:56	SMPL
M2400-07	X3371	25647	6.5073 -		SFAM01.1	PPB	5/28/2021 12:58	SMPL
M2400-08	X3371D	25781	6.5413 -		SFAM01.1	PPB	5/28/2021 13:03	SMPL
M2400-09	X3371S	29593	7.5086 -		SFAM01.1	PPB	5/28/2021 13:07	SMPL
CCV807	CCV087	19411	4.9249 -		SFAM01.1	PPB	5/28/2021 13:10	SMPL
CCB087	CCB087	-65	-0.0173 -		SFAM01.1	PPB	5/28/2021 13:12	SMPL