


LB114693

SFAM01.1

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.00	0.000	0.012	0.012	15	0.000	15					
0.05	0.050					204					
0.20	0.200	0.227	0.027	941	0.0 %	941					14
2.50	2.500	2.459	-0.041	10569	0.0 %	10569					-2
5.00	5.000	5.031	0.031	21665	0.0 %	21665					1
7.50	7.500	7.414	-0.086	31945	0.0 %	31945					1
10.0	10.000	10.058	0.058	43352	0.0 %	43352					1

LB114693

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	15	-		0 SFAM01.1	PPB	5/28/2021 13:58	Std
	0.2 S01	941	-		0.2 SFAM01.1	PPB	5/28/2021 14:02	Std
	2.5 S02	10569	-		2.5 SFAM01.1	PPB	5/28/2021 14:05	Std
	5 S03	21665	-		5 SFAM01.1	PPB	5/28/2021 14:10	Std
	7.5 S04	31945	-		7.5 SFAM01.1	PPB	5/28/2021 14:12	Std
	10 S05	43352	-		10 SFAM01.1	PPB	5/28/2021 14:17	Std
ICV037	ICV037	16953	3.9385	-	SFAM01.1	PPB	5/28/2021 14:24	SMPL
ICB037	ICB037	-80	-0.0098	-	SFAM01.1	PPB	5/28/2021 14:26	SMPL
CCV088	CCV088	21261	4.9371	-	SFAM01.1	PPB	5/28/2021 14:28	SMPL
CCB088	CCB088	-113	-0.0174	-	SFAM01.1	PPB	5/28/2021 14:31	SMPL
M2456-01	MA0BC6	609	0.15	-	SFAM01.1	PPB	5/28/2021 14:35	SMPL
M2456-02	MA0BE2	515	0.1282	-	SFAM01.1	PPB	5/28/2021 14:37	SMPL
M2456-03	MA0BE4	556	0.1377	-	SFAM01.1	PPB	5/28/2021 14:40	SMPL
M2456-04	MA0BE6	499	0.1245	-	SFAM01.1	PPB	5/28/2021 14:42	SMPL
M2456-05	MA0BD6	457	0.1147	-	SFAM01.1	PPB	5/28/2021 14:44	SMPL
M2456-06	MA0BD8	473	0.1184	-	SFAM01.1	PPB	5/28/2021 14:46	SMPL
M2456-07	MA0BE0	507	0.1263	-	SFAM01.1	PPB	5/28/2021 14:49	SMPL
M2456-08	MA0BB7	387	0.0985	-	SFAM01.1	PPB	5/28/2021 14:51	SMPL
M2456-09	MA0BC8	502	0.1252	-	SFAM01.1	PPB	5/28/2021 14:53	SMPL
M2456-10	MA0BD0	527	0.131	-	SFAM01.1	PPB	5/28/2021 14:56	SMPL
M2456-11	MA0BD2	538	0.1335	-	SFAM01.1	PPB	5/28/2021 14:58	SMPL
M2456-12	MA0BD4	216	0.0589	-	SFAM01.1	PPB	5/28/2021 15:00	SMPL
M2456-13	MA0BD4D	210	0.0575	-	SFAM01.1	PPB	5/28/2021 15:02	SMPL
M2456-14	MA0BD4S	4646	1.0857	-	SFAM01.1	PPB	5/28/2021 15:05	SMPL
M2456-15	MA0BG1	18607	4.3219	-	SFAM01.1	PPB	5/28/2021 15:07	SMPL
M2456-16	MA0BB9	15	0.0123	-	SFAM01.1	PPB	5/28/2021 15:09	SMPL
M2456-17	MA0BC3	108	0.0338	-	SFAM01.1	PPB	5/28/2021 15:11	SMPL
M2456-18	MA0BC4	137	0.0405	-	SFAM01.1	PPB	5/28/2021 15:14	SMPL
PB136723BL	PBW723	23	0.0141	-	SFAM01.1	PPB	5/28/2021 15:16	SMPL
CCV089	CCV089	21191	4.9209	-	SFAM01.1	PPB	5/28/2021 15:18	SMPL
CCB089	CCB089	-75	-0.0086	-	SFAM01.1	PPB	5/28/2021 15:20	SMPL
M2458-01	MA0BC7	203	0.0558	-	SFAM01.1	PPB	5/28/2021 15:23	SMPL
M2458-02	MA0BE3	177	0.0498	-	SFAM01.1	PPB	5/28/2021 15:25	SMPL
M2458-03	MA0BE5	190	0.0528	-	SFAM01.1	PPB	5/28/2021 15:27	SMPL
M2458-04	MA0BE7	162	0.0463	-	SFAM01.1	PPB	5/28/2021 15:29	SMPL
M2458-05	MA0BD7	197	0.0545	-	SFAM01.1	PPB	5/28/2021 15:32	SMPL
M2458-06	MA0BD9	189	0.0526	-	SFAM01.1	PPB	5/28/2021 15:34	SMPL
M2458-07	MA0BE1	210	0.0575	-	SFAM01.1	PPB	5/28/2021 15:36	SMPL
M2458-08	MA0BB8	163	0.0466	-	SFAM01.1	PPB	5/28/2021 15:39	SMPL
M2458-09	MA0BC9	178	0.0501	-	SFAM01.1	PPB	5/28/2021 15:41	SMPL
M2458-10	MA0BD1	192	0.0533	-	SFAM01.1	PPB	5/28/2021 15:43	SMPL
M2458-11	MA0BD3	208	0.057	-	SFAM01.1	PPB	5/28/2021 15:45	SMPL
M2458-12	MA0BD5	759	0.1847	-	SFAM01.1	PPB	5/28/2021 15:48	SMPL
M2458-13	MA0BD5D	736	0.1794	-	SFAM01.1	PPB	5/28/2021 15:50	SMPL
M2458-14	MA0BD5S	4839	1.1305	-	SFAM01.1	PPB	5/28/2021 15:52	SMPL
M2458-15	MA0BG2	17680	4.107	-	SFAM01.1	PPB	5/28/2021 15:54	SMPL

LB114693

INSTRUMENT ID : CV1

M2458-16	MA0BC0	335	0.0864 -	SFAM01.1	PPB	5/28/2021 15:57 SMPL
PB136722BL	PBW722	135	0.0401 -	SFAM01.1	PPB	5/28/2021 16:14 SMPL
CCV090	CCV090	20760	4.821 -	SFAM01.1	PPB	5/28/2021 16:17 SMPL
CCB090	CCB090	-124	-0.02 -	SFAM01.1	PPB	5/28/2021 16:19 SMPL