



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB114700

Slope : 97.9

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2718
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2809
buffer solution pH 7 yellow	W2719
Buffer Solution, PH2 (500ml)	W2806
Buffer Solution, PH12 (500ml)	W2807

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	05/28/2021	15:20
2	CAL2	1	Water	NA	NA	20.3	7.00	05/28/2021	15:21
3	CAL3	1	Water	NA	NA	20.3	10.02	05/28/2021	15:22
4	ICV	1	Water	NA	NA	20.4	6.99	05/28/2021	15:24
5	CCV1	1	Water	NA	NA	20.4	2.01	05/28/2021	15:25
6	M2569-05	1	Water	NA	NA	23.1	9.28	05/28/2021	15:36
7	M2569-05DUP	1	Water	NA	NA	23.2	9.29	05/28/2021	15:37
8	CCV2	1	Water	NA	NA	20.2	12.02	05/28/2021	15:39

WORKLIST(Hardcopy Internal Chain)

28114700

WorkList Name : CORROSIONITY M2569 V WorkList ID : 149356 Department : Wet-Chemistry Date : 05-28-2021 15:12:33

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
05/27/2021	Water	M2569-05	Corrosivity	Cool 4 deg C	TETR16	D62	MC0AA1	05/27/2021	9040C

Date/Time 05/28/21 15:15
 Raw Sample Received by: JD Post (JL Post)
 Raw Sample Relinquished by: CP Sm

Date/Time 05/28/21 16:30
 Raw Sample Received by: CP Sm
 Raw Sample Relinquished by: JD at Chen