


LB114701

7470A

INSTRUMENT ID: CV2

 μ Abs.:

36677



Linear

A= 0.0000e+000

B= 2.7368e-004

C= -6.8740e-003

Rho= 0.9999768

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.005	0.005	42	0.000	42					0/0/1
0.2	0.200	0.194	-0.006	733	0.0 %	733					1
2.5	2.500	2.516	0.016	9218	0.0 %	9218					3
5.0	5.000	5.005	0.005	18313	0.0 %	18313					1
7.5	7.500	7.450	-0.050	27247	0.0 %	27247					0
10.0	10.000	10.031	0.031	36677	0.0 %	36677					0

LB114701 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 01 Jun 2021 09:10:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S ₀	42	-	0.0000	7470A	PPB	01 Jun 2021 09:30:50	S	Std
0.2 - 1	S ₀₋₂	733	-	0.2000	7470A	PPB	01 Jun 2021 09:33:07	S	Std
2.5 - 1	S ₂₋₅	9218	-	2.5000	7470A	PPB	01 Jun 2021 09:35:24	S	Std
5.0 - 1	S ₅	18313	-	5.0000	7470A	PPB	01 Jun 2021 09:37:41	S	Std
7.5 - 1	S ₇₋₅	27247	-	7.5000	7470A	PPB	01 Jun 2021 09:39:57	S	Std
10.0 - 1	S ₁₀	36677	-	10.0000	7470A	PPB	01 Jun 2021 09:42:13	S	Std
ICV42 - 1	ICV42	14693	4.0143	-7470A	PPB	01 Jun 2021 09:45:00	U	SMPL	
ICB42 - 1	ICB42	-163	-0.0515	-7470A	PPB	01 Jun 2021 09:47:15	U	SMPL	
CCV91 - 1	CCV91	17286	4.7239	-7470A	PPB	01 Jun 2021 09:49:30	U	SMPL	
CCB91 - 1	CCB91	-237	-0.0717	-7470A	PPB	01 Jun 2021 09:51:45	U	SMPL	
CRA - 1	CRA	617	0.1620	-7470A	PPB	01 Jun 2021 09:54:02	U	SMPL	
HighStd - 1	HighStd	34302	9.3808	-7470A	PPB	01 Jun 2021 09:56:16	U	SMPL	
ChkStd - 1	ChkStd	24589	6.7226	-7470A	PPB	01 Jun 2021 10:01:13	U	SMPL	
PB136728BL - 1	PBW	-48	-0.0200	-7470A	PPB	01 Jun 2021 10:06:03	U	SMPL	
PB136728BS - 1	LCSW	13002	3.5515	-7470A	PPB	01 Jun 2021 10:08:19	U	SMPL	
M2530-02 - 1	VB15131	-205	-0.0630	-7470A	PPB	01 Jun 2021 10:10:35	U	SMPL	
M2551-02 - 1	JCSY-SP1	4	-0.0058	-7470A	PPB	01 Jun 2021 10:12:52	U	SMPL	
M2551-06 - 1	JCSY-SP2	18	-0.0019	-7470A	PPB	01 Jun 2021 10:15:09	U	SMPL	
M2551-10 - 1	JCSY-SP3	25	-0.0000	-7470A	PPB	01 Jun 2021 10:17:26	U	SMPL	
M2560-02 - 1	S1-COMP-A-B-C-D	37	0.0033	-7470A	PPB	01 Jun 2021 10:19:42	U	SMPL	
CCV92 - 1	CCV92	17038	4.6560	-7470A	PPB	01 Jun 2021 10:34:01	U	SMPL	
CCB92 - 1	CCB92	-125	-0.0411	-7470A	PPB	01 Jun 2021 10:36:16	U	SMPL	
M2560-02DUP - 1	S1-COMP-A-B-C-DDUP	-2	-0.0074	-7470A	PPB	01 Jun 2021 10:38:32	U	SMPL	
M2560-02MS - 1	S1-COMP-A-B-C-DMS	13241	3.6169	-7470A	PPB	01 Jun 2021 10:43:35	U	SMPL	
M2560-02MSD - 1	S1-COMP-A-B-C-DMSD	13336	3.6429	-7470A	PPB	01 Jun 2021 10:45:50	U	SMPL	
PB136729BL - 1	PBW	-71	-0.0263	-7470A	PPB	01 Jun 2021 10:48:06	U	SMPL	
PB136729BS - 1	LCSW	13231	3.6141	-7470A	PPB	01 Jun 2021 10:50:22	U	SMPL	
M2530-04 - 1	VB14058	379	0.0968	-7470A	PPB	01 Jun 2021 10:52:38	U	SMPL	
M2556-01 - 1	MW-8B	2	-0.0063	-7470A	PPB	01 Jun 2021 10:54:55	U	SMPL	
M2556-01DUP - 1	MW-8BDUP	15	-0.0028	-7470A	PPB	01 Jun 2021 10:57:11	U	SMPL	
M2556-01MS - 1	MW-8BMS	13202	3.6062	-7470A	PPB	01 Jun 2021 10:59:28	U	SMPL	
M2556-01MSD - 1	MW-8BMSD	13411	3.6634	-7470A	PPB	01 Jun 2021 11:01:44	U	SMPL	
CCV93 - 1	CCV93	17621	4.8156	-7470A	PPB	01 Jun 2021 11:03:58	U	SMPL	
CCB93 - 1	CCB93	-70	-0.0260	-7470A	PPB	01 Jun 2021 11:06:14	U	SMPL	
M2556-02 - 1	MW-5B	36	0.0030	-7470A	PPB	01 Jun 2021 11:08:30	U	SMPL	
M2512-02 - 1	COMP-1-3	30	0.0013	-7470A	PPB	01 Jun 2021 11:10:45	U	SMPL	
M2512-05 - 1	COMP-4-6	24	-0.0003	-7470A	PPB	01 Jun 2021 11:13:01	U	SMPL	
PB136728TB - 1	PB136696TB	28	0.0008	-7470A	PPB	01 Jun 2021 11:15:16	U	SMPL	
PB136729TB - 1	PB136708TB	38	0.0035	-7470A	PPB	01 Jun 2021 11:17:33	U	SMPL	
M2560-02LX5 - 1		34	0.0024	-7470A	PPB	01 Jun 2021 11:19:49	U	SMPL	
M2560-02A - 1		13940	3.8082	-7470A	PPB	01 Jun 2021 11:22:05	U	SMPL	
M2556-01LX5 - 1		-24	-0.0134	-7470A	PPB	01 Jun 2021 11:24:22	U	SMPL	
M2556-01A - 1		14392	3.9319	-7470A	PPB	01 Jun 2021 11:31:31	U	SMPL	
CCV94 - 1	CCV94	18396	5.0277	-7470A	PPB	01 Jun 2021 11:33:46	U	SMPL	
CCB94 - 1	CCB94	-95	-0.0329	-7470A	PPB	01 Jun 2021 11:36:02	U	SMPL	