



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114717

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	06/01/2021	09:55
2	M2564-08		1	8	100.00	>212.0	>212.0	06/01/2021	10:25
3	M2564-08DUP		1	8	100.00	>212.0	>212.0	06/01/2021	10:45
4	M2564-10		1	9	100.00	>212.0	>212.0	06/01/2021	11:10
5	M2569-05		1	9	100.00	>212.0	>212.0	06/01/2021	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114717

WorkList Name : FP6121 WorkList ID : 149390 Department : Wet-Chemistry Date : 06-01-2021 09:24:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/04/2021	Water	M2564-08	C Flash Point	Cool 4 deg C	PSEG03	E61	33211	05/28/2021	1010A
06/04/2021	Water	M2564-10	C Flash Point	Cool 4 deg C	PSEG03	E61	33212	05/28/2021	1010A
06/04/2021	Water	M2569-05	D Flash Point	Cool 4 deg C	TETR16	D62	MC0AA1	05/27/2021	1010A

Date/Time 06/01/21 09:30
Raw Sample Received by: KS (Webster)
Raw Sample Relinquished by: KS (Webster)

Date/Time 06/01/21 11:40
Raw Sample Received by: KS (Webster)
Raw Sample Relinquished by: KS (Webster)