



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB114718

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.3	06/01/2021	09:55
2	M2564-01		1	8	100.00	>212.0	>212.0	06/01/2021	11:58
3	M2564-01DUP		1	8	100.00	>212.0	>212.0	06/01/2021	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB114718

WorkList Name : FP6121-1 WorkList ID : 149391 Department : Wet-Chemistry Date : 06-01-2021 09:25:49

Due Date	Matrix	Sample	Test	Preservative	Customer	Raw Sample Storage Location	Customer Sample	Collect Date	Method
06/04/2021	Solid	M2564-01	✓ Flash Point	Cool 4 deg C	PSEG03	E61	33211-OIL	05/28/2021	1010A

Date/Time 06/01/21 09:30
Raw Sample Received by: KS (Wickher)
Raw Sample Relinquished by: KS (Wickher)