


245.1

LB114720

INSTRUMENT ID: CV1

 μ Abs.:

34517



Linear

A= 0.0000e+000

B= 2.8778e-004

C= -1.3968e-002

Rho= 0.9999025

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/100
0.0	0.000	-0.007	-0.007	25	0.000	25					-
0.05	0.050	0.050	-0.000	222	0.0 %	222	0				0
0.2	0.200	0.199	-0.001	741	0.0 %	741					-1
2.5	2.500	2.463	-0.037	8608	0.0 %	8608					-1
5.0	5.000	5.023	0.023	17503	0.0 %	17503					0
7.5	7.500	7.602	0.102	26464	0.0 %	26464					1
10.0	10.000	9.919	-0.081	34517	0.0 %	34517					-1

LB114724 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 01 Jun 2021 14:03:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	25	-	0.0000	245.1	PPB	01 Jun 2021 14:05:18	S	Std
0.05 - 1	SD.05	222	-	0.0500	245.1	PPB	01 Jun 2021 14:07:35	S	Std
0.2 - 1	SD.2	741	-	0.2000	245.1	PPB	01 Jun 2021 14:09:52	S	Std
2.5 - 1	SD.5	8608	-	2.5000	245.1	PPB	01 Jun 2021 14:12:09	S	Std
5.0 - 1	SD	17503	-	5.0000	245.1	PPB	01 Jun 2021 14:14:25	S	Std
7.5 - 1	SD	26464	-	7.5000	245.1	PPB	01 Jun 2021 14:16:40	S	Std
10.0 - 1	SD	34517	-	10.0000	245.1	PPB	01 Jun 2021 14:18:56	S	Std
ICV43 - 1	ICV43	13972	4.0069	-245.1		PPB	01 Jun 2021 14:21:50	U	SMPL
ICB43 - 1	ICB43	-38	-0.0249	-245.1		PPB	01 Jun 2021 14:24:04	U	SMPL
CCV95 - 1	CCV95	17630	5.0596	-245.1		PPB	01 Jun 2021 14:26:19	U	SMPL
CCB95 - 1	CCB95	-52	-0.0289	-245.1		PPB	01 Jun 2021 14:28:34	U	SMPL
CRA - 1	CRA	224	0.0505	-245.1		PPB	01 Jun 2021 14:30:49	U	SMPL
HighStd - 1	HighStd	34582	9.9381	-245.1		PPB	01 Jun 2021 14:33:04	U	SMPL
ChkStd - 1	ChkStd	25486	7.3204	-245.1		PPB	01 Jun 2021 14:35:19	U	SMPL
PB136752BL - 1	PBW	-84	-0.0381	-245.1		PPB	01 Jun 2021 14:37:35	U	SMPL
PB136752BS - 1	LCSW	13371	3.8340	-245.1		PPB	01 Jun 2021 14:39:51	U	SMPL
M2532-01 - 1	MH-N-(MH-10.29)	11	-0.0108	-245.1		PPB	01 Jun 2021 14:42:07	U	SMPL
M2532-01DUP - 1	MH-N-(MH-10.29)DUP	56	0.0021	-245.1		PPB	01 Jun 2021 14:44:23	U	SMPL
M2532-01MS - 1	MH-N-(MH-10.29)MS	14947	4.2875	-245.1		PPB	01 Jun 2021 14:46:40	U	SMPL
M2532-01MSD - 1	MH-N-(MH-10.29)MSD	15328	4.3971	-245.1		PPB	01 Jun 2021 14:48:59	U	SMPL
M2532-02 - 1	MH-N-(MH-10.29)	40	-0.0025	-245.1		PPB	01 Jun 2021 14:51:14	U	SMPL
CCV96 - 1	CCV96	17529	5.0306	-245.1		PPB	01 Jun 2021 14:53:30	U	SMPL
CCB96 - 1	CCB96	-46	-0.0272	-245.1		PPB	01 Jun 2021 14:55:46	U	SMPL
M2533-01 - 1	DEMMON-WATER-TANK	37	-0.0033	-245.1		PPB	01 Jun 2021 14:58:01	U	SMPL
M2532-01LX5 - 1		38	-0.0030	-245.1		PPB	01 Jun 2021 15:00:16	U	SMPL
M2532-01A - 1		14708	4.2187	-245.1		PPB	01 Jun 2021 15:02:32	U	SMPL
CCV97 - 1	CCV97	17689	5.0766	-245.1		PPB	01 Jun 2021 15:04:47	U	SMPL
CCB97 - 1	CCB97	-40	-0.0255	-245.1		PPB	01 Jun 2021 15:07:02	U	SMPL